



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2372

Contract: GENV01

Lab Code: CHEM

Case No.: Q2372

SAS No.: Q2372

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	12:30	LB136227

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2372

Contract: GENV01

Lab Code: CHEM

Case No.: Q2372

SAS No.: Q2372

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	12:35	LB136227
CCB02	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	13:05	LB136227
CCB03	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	13:32	LB136227
CCB04	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	13:53	LB136227
CCB05	Mercury	0.076	+/-0.2	U	0.20	CV	06/23/2025	14:22	LB136227

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	14.8	+/-50	J	100	P	06/24/2025	19:08	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	19:08	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	19:08	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	19:08	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	19:08	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	19:08	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	19:08	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	19:08	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	19:08	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	19:08	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	19:08	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	19:08	LB136255
	Magnesium	244	+/-1000	U	2000	P	06/24/2025	19:08	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	19:08	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	19:08	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	19:08	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	19:08	LB136255
	Silver	1.62	+/-5	U	10.0	P	06/24/2025	19:08	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	19:08	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	19:08	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	19:08	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	19:08	LB136255

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/24/2025	19:39	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	19:39	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	19:39	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	19:39	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	19:39	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	19:39	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	19:39	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	19:39	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	19:39	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	19:39	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	19:39	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	19:39	LB136255
	Magnesium	244	+/-1000	U	2000	P	06/24/2025	19:39	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	19:39	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	19:39	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	19:39	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	19:39	LB136255
	Silver	1.62	+/-5	U	10.0	P	06/24/2025	19:39	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	19:39	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	19:39	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	19:39	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	19:39	LB136255
CCB02	Aluminum	11.3	+/-50	U	100	P	06/24/2025	20:41	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	20:41	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	20:41	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	20:41	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	20:41	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	20:41	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	20:41	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	20:41	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	20:41	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	20:41	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	20:41	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	20:41	LB136255
	Magnesium	244	+/-1000	U	2000	P	06/24/2025	20:41	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	20:41	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	20:41	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	20:41	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	20:41	LB136255

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/24/2025	20:41	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	20:41	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	20:41	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	20:41	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	20:41	LB136255
CCB03	Aluminum	11.3	+/-50	U	100	P	06/24/2025	21:33	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	21:33	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	21:33	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	21:33	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	21:33	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	21:33	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	21:33	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	21:33	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	21:33	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	21:33	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	21:33	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	21:33	LB136255
	Magnesium	244	+/-1000	U	2000	P	06/24/2025	21:33	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	21:33	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	21:33	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	21:33	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	21:33	LB136255
	Silver	1.62	+/-5	U	10.0	P	06/24/2025	21:33	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	21:33	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	21:33	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	21:33	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	21:33	LB136255
CCB04	Aluminum	11.3	+/-50	U	100	P	06/24/2025	23:01	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	23:01	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	23:01	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	23:01	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	23:01	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	23:01	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	23:01	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	23:01	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	23:01	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	23:01	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	23:01	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	23:01	LB136255

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2372							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2372	SAS No.: Q2372						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/24/2025	23:01	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	23:01	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	23:01	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	23:01	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	23:01	LB136255
	Silver	1.62	+/-5	U	10.0	P	06/24/2025	23:01	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	23:01	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	23:01	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	23:01	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	23:01	LB136255
CCB05	Aluminum	11.3	+/-50	U	100	P	06/24/2025	23:50	LB136255
	Antimony	6.76	+/-25	U	50.0	P	06/24/2025	23:50	LB136255
	Arsenic	5.12	+/-10	U	20.0	P	06/24/2025	23:50	LB136255
	Barium	14.6	+/-50	U	100	P	06/24/2025	23:50	LB136255
	Beryllium	0.56	+/-3	U	6.00	P	06/24/2025	23:50	LB136255
	Cadmium	0.50	+/-3	U	6.00	P	06/24/2025	23:50	LB136255
	Calcium	234	+/-1000	U	2000	P	06/24/2025	23:50	LB136255
	Chromium	2.12	+/-5	U	10.0	P	06/24/2025	23:50	LB136255
	Cobalt	2.26	+/-15	U	30.0	P	06/24/2025	23:50	LB136255
	Copper	4.60	+/-10	U	20.0	P	06/24/2025	23:50	LB136255
	Iron	23.4	+/-50	U	100	P	06/24/2025	23:50	LB136255
	Lead	2.30	+/-6	U	12.0	P	06/24/2025	23:50	LB136255
	Magnesium	244	+/-1000	U	2000	P	06/24/2025	23:50	LB136255
	Manganese	5.94	+/-10	U	20.0	P	06/24/2025	23:50	LB136255
	Nickel	3.06	+/-20	U	40.0	P	06/24/2025	23:50	LB136255
	Potassium	918	+/-1000	U	2000	P	06/24/2025	23:50	LB136255
	Selenium	9.64	+/-10	U	20.0	P	06/24/2025	23:50	LB136255
	Silver	1.62	+/-5	U	10.0	P	06/24/2025	23:50	LB136255
	Sodium	868	+/-1000	U	2000	P	06/24/2025	23:50	LB136255
	Thallium	4.38	+/-20	U	40.0	P	06/24/2025	23:50	LB136255
	Vanadium	6.26	+/-20	U	40.0	P	06/24/2025	23:50	LB136255
	Zinc	3.50	+/-20	U	40.0	P	06/24/2025	23:50	LB136255

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/25/2025	14:53	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	14:53	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	14:53	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	14:53	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	14:53	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	14:53	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	14:53	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	14:53	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	14:53	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	14:53	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	14:53	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	14:53	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	14:53	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	14:53	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	14:53	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	14:53	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	14:53	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	14:53	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	14:53	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	14:53	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	14:53	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	14:53	LB136273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/25/2025	16:09	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	16:09	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	16:09	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	16:09	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	16:09	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	16:09	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	16:09	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	16:09	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	16:09	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	16:09	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	16:09	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	16:09	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	16:09	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	16:09	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	16:09	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	16:09	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	16:09	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	16:09	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	16:09	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	16:09	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	16:09	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	16:09	LB136273
CCB02	Aluminum	11.3	+/-50	U	100	P	06/25/2025	17:09	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	17:09	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	17:09	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	17:09	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	17:09	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	17:09	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	17:09	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	17:09	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	17:09	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	17:09	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	17:09	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	17:09	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	17:09	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	17:09	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	17:09	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	17:09	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	17:09	LB136273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2372							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2372	SAS No.: Q2372						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/25/2025	17:09	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	17:09	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	17:09	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	17:09	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	17:09	LB136273
CCB03	Aluminum	22.1	+/-50	J	100	P	06/25/2025	18:07	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	18:07	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	18:07	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	18:07	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	18:07	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	18:07	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	18:07	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	18:07	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	18:07	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	18:07	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	18:07	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	18:07	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	18:07	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	18:07	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	18:07	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	18:07	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	18:07	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	18:07	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	18:07	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	18:07	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	18:07	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	18:07	LB136273
CCB04	Aluminum	11.3	+/-50	U	100	P	06/25/2025	18:59	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	18:59	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	18:59	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	18:59	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	18:59	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	18:59	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	18:59	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	18:59	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	18:59	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	18:59	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	18:59	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	18:59	LB136273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2372							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2372	SAS No.: Q2372						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/25/2025	18:59	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	18:59	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	18:59	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	18:59	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	18:59	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	18:59	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	18:59	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	18:59	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	18:59	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	18:59	LB136273
CCB05	Aluminum	11.3	+/-50	U	100	P	06/25/2025	19:57	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	19:57	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	19:57	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	19:57	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	19:57	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	19:57	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	19:57	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	19:57	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	19:57	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	19:57	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	19:57	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	19:57	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	19:57	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	19:57	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	19:57	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	19:57	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	19:57	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	19:57	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	19:57	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	19:57	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	19:57	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	19:57	LB136273
CCB06	Aluminum	11.3	+/-50	U	100	P	06/25/2025	20:50	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	20:50	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	20:50	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	20:50	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	20:50	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	20:50	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	20:50	LB136273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2372</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2372</u>	SAS No.: <u>Q2372</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	9.17	+/-5	J*	10.0	P	06/25/2025	20:50	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	20:50	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	20:50	LB136273
	Iron	33.2	+/-50	J	100	P	06/25/2025	20:50	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	20:50	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	20:50	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	20:50	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	20:50	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	20:50	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	20:50	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	20:50	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	20:50	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	20:50	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	20:50	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	20:50	LB136273
CCB07	Aluminum	11.3	+/-50	U	100	P	06/25/2025	21:16	LB136273
	Antimony	6.76	+/-25	U	50.0	P	06/25/2025	21:16	LB136273
	Arsenic	5.12	+/-10	U	20.0	P	06/25/2025	21:16	LB136273
	Barium	14.6	+/-50	U	100	P	06/25/2025	21:16	LB136273
	Beryllium	0.56	+/-3	U	6.00	P	06/25/2025	21:16	LB136273
	Cadmium	0.50	+/-3	U	6.00	P	06/25/2025	21:16	LB136273
	Calcium	234	+/-1000	U	2000	P	06/25/2025	21:16	LB136273
	Chromium	2.12	+/-5	U	10.0	P	06/25/2025	21:16	LB136273
	Cobalt	2.26	+/-15	U	30.0	P	06/25/2025	21:16	LB136273
	Copper	4.60	+/-10	U	20.0	P	06/25/2025	21:16	LB136273
	Iron	23.4	+/-50	U	100	P	06/25/2025	21:16	LB136273
	Lead	2.30	+/-6	U	12.0	P	06/25/2025	21:16	LB136273
	Magnesium	244	+/-1000	U	2000	P	06/25/2025	21:16	LB136273
	Manganese	5.94	+/-10	U	20.0	P	06/25/2025	21:16	LB136273
	Nickel	3.06	+/-20	U	40.0	P	06/25/2025	21:16	LB136273
	Potassium	918	+/-1000	U	2000	P	06/25/2025	21:16	LB136273
	Selenium	9.64	+/-10	U	20.0	P	06/25/2025	21:16	LB136273
	Silver	1.62	+/-5	U	10.0	P	06/25/2025	21:16	LB136273
	Sodium	868	+/-1000	U	2000	P	06/25/2025	21:16	LB136273
	Thallium	4.38	+/-20	U	40.0	P	06/25/2025	21:16	LB136273
	Vanadium	6.26	+/-20	U	40.0	P	06/25/2025	21:16	LB136273
	Zinc	3.50	+/-20	U	40.0	P	06/25/2025	21:16	LB136273