

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2372</u>
Contract:	<u>GENV01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2372</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2372-02</u>	Client ID:	<u>GAV2WDUP</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>Q2372-02DUP</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.71		2.70		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2372</u>
Contract:	<u>GENV01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2372</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2372-02MS</u>	Client ID:	<u>GAV2WMSD</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>Q2372-02MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	5.28		5.34		1		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: G Environmental **Level:** LOW **SDG No.:** Q2372
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2372 **SAS No.:** Q2372
Matrix: Water **Sample ID:** Q2385-01 **Client ID:** A5311DUP
Percent Solids for Sample: NA **Duplicate ID** Q2385-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	63.7		62.9		1		P
Antimony	ug/L	20	3.62	J	3.43	J	5		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	56.7		57.3		1		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	33300		33600		1		P
Chromium	ug/L	20	5.00	U	5.00	U			P
Cobalt	ug/L	20	7.25	J	7.16	J	1		P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	286		289		1		P
Lead	ug/L	20	3.23	J	3.36	J	4		P
Magnesium	ug/L	20	15600		15500		1		P
Manganese	ug/L	20	28.1		29.1		3		P
Nickel	ug/L	20	20.0	U	20.0	U			P
Potassium	ug/L	20	1810		1830		1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	44200		44600		1		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	4.12	J	4.29	J	4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: G Environmental **Level:** LOW **SDG No.:** Q2372
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2372 **SAS No.:** Q2372
Matrix: Water **Sample ID:** Q2385-01MS **Client ID:** A5311MSD
Percent Solids for Sample: NA **Duplicate ID** Q2385-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1110		1060		5		P
Antimony	ug/L	20	421		413		2		P
Arsenic	ug/L	20	390		383		2		P
Barium	ug/L	20	161		158		2		P
Beryllium	ug/L	20	108		107		1		P
Cadmium	ug/L	20	99.8		97.6		2		P
Calcium	ug/L	20	34100		33900		1		P
Chromium	ug/L	20	212		209		1		P
Cobalt	ug/L	20	108		106		2		P
Copper	ug/L	20	162		159		2		P
Iron	ug/L	20	1910		1880		2		P
Lead	ug/L	20	489		478		2		P
Magnesium	ug/L	20	16800		16500		2		P
Manganese	ug/L	20	133		131		2		P
Nickel	ug/L	20	250		244		2		P
Potassium	ug/L	20	7520		7410		1		P
Selenium	ug/L	20	982		960		2		P
Silver	ug/L	20	34.6		29.7		15		P
Sodium	ug/L	20	45800		45400		1		P
Thallium	ug/L	20	971		952		2		P
Vanadium	ug/L	20	159		158		1		P
Zinc	ug/L	20	109		108		1		P