

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2372

Contract: GENV01

Lab Code: CHEM

Case No.: Q2372

SAS No.: Q2372

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168568BS Mercury	ug/L	4.0	3.76		94	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q2372
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2372 **SAS No.:** Q2372

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168583BS							
Aluminum	ug/L	1000	1010		101	80 - 120	P
Antimony	ug/L	400	411		103	80 - 120	P
Arsenic	ug/L	400	364		91	80 - 120	P
Barium	ug/L	100	101		101	80 - 120	P
Beryllium	ug/L	100	106		106	80 - 120	P
Cadmium	ug/L	100	96.5		96	80 - 120	P
Calcium	ug/L	500	493	J	99	80 - 120	P
Chromium	ug/L	200	201		100	80 - 120	P
Cobalt	ug/L	100	98.5		98	80 - 120	P
Copper	ug/L	150	159		106	80 - 120	P
Iron	ug/L	1500	1570		105	80 - 120	P
Lead	ug/L	500	481		96	80 - 120	P
Magnesium	ug/L	1000	1030		103	80 - 120	P
Manganese	ug/L	100	103		103	80 - 120	P
Nickel	ug/L	250	248		99	80 - 120	P
Potassium	ug/L	5000	4140		83	80 - 120	P
Selenium	ug/L	1000	1020		102	80 - 120	P
Silver	ug/L	37.5	38.2		102	80 - 120	P
Sodium	ug/L	1500	1420		95	80 - 120	P
Thallium	ug/L	1000	1010		101	80 - 120	P
Vanadium	ug/L	150	149		99	80 - 120	P
Zinc	ug/L	100	104		104	80 - 120	P