

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG No.: Q2372

Instrument ID: BNA_N Calibration Date/Time: 06/27/2025 10:09

Lab File ID: BN037417.D Init. Calib. Date(s): 06/26/2025 06/26/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 10:41 14:17

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.619	0.578		-6.6	20.0
Fluoranthene-d10	1.146	1.126		-1.7	20.0
2-Fluorophenol	0.771	0.793		2.9	20.0
Phenol-d6	0.799	0.783		-2.0	20.0
Nitrobenzene-d5	0.321	0.297		-7.5	20.0
2-Fluorobiphenyl	1.727	1.669		-3.4	20.0
2,4,6-Tribromophenol	0.235	0.245		4.3	20.0
Terphenyl-d14	0.853	0.854		0.1	20.0
Benzo(a)anthracene	1.257	1.166		-7.2	20.0
Benzo(b)fluoranthene	1.410	1.335		-5.3	20.0
Benzo(k)fluoranthene	1.507	1.509		0.1	20.0
Benzo(a)pyrene	1.254	1.180		-5.9	20.0
Benzo(g,h,i)perylene	1.548	1.484		-4.1	20.0

All other compounds must meet a minimum RRF of 0.010.