

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG NO.: Q2372

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/27/2025

Lab File ID: BN037417.D Time Analyzed: 10:09

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6217	16.959	6164	21.162	6622	23.345
UPPER LIMIT	12434	17.459	12328	21.662	13244	23.845
LOWER LIMIT	3108.5	16.459	3082	20.662	3311	22.845
EPA SAMPLE NO.						
01 GAV1WRE	10583	16.96	13030 *	21.16	12884	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.