

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q2372

SDG NO.: Q2372

Lab File ID: BN037402.D

DFTPP Injection Date: 06/26/2025

Instrument ID: BNA_N

DFTPP Injection Time: 20:19

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.3
365	Greater than 1% of mass 198	6.1
441	Present, but less than mass 443	83.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	13.7 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037403.D	06/26/2025	21:38
GAV1W	Q2372-01	BN037405.D	06/26/2025	22:50