

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG No.: Q2372
Instrument ID: BNA_F Calibration Date(s): 06/19/2025 06/19/2025
Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.238	1.230	1.230	1.149	1.263	1.201	4.3
Benzaldehyde			0.988	0.943	0.757	0.832	0.841	15.0
Phenol-d6		1.512	1.481	1.423	1.340	1.480	1.417	5.5
bis(2-Chloroethyl)ether		1.167	1.133	1.134	1.072	1.183	1.126	3.7
2,2-oxybis(1-Chloropropane)		1.975	1.869	1.815	1.723	1.879	1.809	6.1
Acetophenone		0.464	0.451	0.459	0.418	0.461	0.441	5.3
n-Nitroso-di-n-propylamine	0.831	0.879	0.840	0.833	0.787	0.847	0.824	4.3
Nitrobenzene-d5		0.369	0.362	0.373	0.354	0.386	0.365	3.5
Hexachloroethane		0.533	0.521	0.507	0.499	0.539	0.511	3.9
Nitrobenzene		0.345	0.325	0.332	0.319	0.349	0.330	3.9
Isophorone		0.611	0.582	0.592	0.555	0.616	0.585	4.2
bis(2-Chloroethoxy)methane		0.386	0.373	0.383	0.355	0.390	0.372	4.1
Naphthalene		1.051	1.002	1.004	0.934	1.017	0.979	5.5
4-Chloroaniline		0.414	0.408	0.409	0.379	0.421	0.398	5.3
Hexachlorobutadiene		0.195	0.188	0.198	0.184	0.200	0.190	4.3
Caprolactam			0.073	0.075	0.072	0.081	0.075	5.8
2-Methylnaphthalene		0.655	0.621	0.629	0.579	0.631	0.607	6.1
Hexachlorocyclopentadiene			0.235	0.308	0.336	0.365	0.335	17.0
2-Fluorobiphenyl		1.740	1.636	1.592	1.430	1.500	1.522	9.2
1,1-Biphenyl		1.701	1.616	1.637	1.505	1.609	1.580	5.3
2-Chloronaphthalene		1.300	1.202	1.208	1.129	1.209	1.186	5.5
2-Nitroaniline		0.321	0.329	0.337	0.324	0.358	0.332	3.9
Dimethylphthalate		1.372	1.334	1.330	1.228	1.335	1.295	5.2
Acenaphthylene		2.110	2.006	2.011	1.861	1.995	1.952	5.6
2,6-Dinitrotoluene		0.297	0.279	0.290	0.277	0.294	0.284	3.6
3-Nitroaniline		0.311	0.311	0.316	0.302	0.338	0.313	4.5
Acenaphthene		1.247	1.199	1.241	1.142	1.228	1.191	4.4
Dibenzofuran		1.844	1.772	1.763	1.626	1.752	1.708	6.1
2,4-Dinitrotoluene		0.373	0.378	0.397	0.367	0.407	0.378	5.9
Diethylphthalate		1.328	1.302	1.319	1.211	1.318	1.269	5.5
4-Chlorophenyl-phenylether		0.713	0.656	0.669	0.599	0.640	0.636	7.6
Fluorene		1.464	1.412	1.395	1.260	1.362	1.334	7.6
4-Nitroaniline		0.279	0.276	0.297	0.272	0.316	0.286	6.5
n-Nitrosodiphenylamine		0.736	0.686	0.710	0.670	0.717	0.693	4.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG No.: Q2372
Instrument ID: BNA_F Calibration Date(s): 06/19/2025 06/19/2025
Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,4,6-Tribromophenol		0.208	0.207	0.214	0.200	0.219	0.206	5.0
4-Bromophenyl-phenylether		0.231	0.221	0.235	0.218	0.236	0.228	3.0
Hexachlorobenzene		0.263	0.248	0.257	0.241	0.262	0.253	3.3
Atrazine		0.176	0.172	0.180	0.174	0.195	0.179	4.5
Phenanthrene		1.183	1.113	1.115	1.031	1.122	1.083	6.2
Anthracene		1.226	1.123	1.149	1.064	1.140	1.111	6.3
Carbazole		1.040	0.986	0.999	0.900	1.002	0.959	6.7
Di-n-butylphthalate		0.997	1.006	1.045	0.946	1.073	0.999	4.9
Fluoranthene		1.136	1.088	1.090	0.954	1.059	1.028	8.4
Pyrene		2.110	1.940	1.968	1.742	2.030	1.875	11.4
Terphenyl-d14		1.691	1.556	1.533	1.327	1.535	1.447	13.3
Butylbenzylphthalate		0.478	0.470	0.538	0.547	0.613	0.544	10.0
3,3-Dichlorobenzidine			0.398	0.440	0.429	0.476	0.438	5.7
Benzo (a) anthracene		1.392	1.358	1.366	1.265	1.403	1.349	4.0
Chrysene		1.243	1.191	1.208	1.164	1.295	1.204	4.2
Bis(2-ethylhexyl)phthalate		0.650	0.675	0.767	0.803	0.861	0.782	11.5
Di-n-octyl phthalate			1.207	1.352	1.463	1.610	1.475	11.5
Benzo (b) fluoranthene		1.164	1.220	1.226	1.072	1.152	1.157	5.6
Benzo (k) fluoranthene		1.206	1.004	1.035	1.037	1.193	1.083	7.5
Benzo (a) pyrene		1.142	1.065	1.124	1.069	1.201	1.118	4.4
Indeno (1,2,3-cd) pyrene		1.528	1.488	1.545	1.474	1.648	1.523	4.4
Dibenzo (a,h) anthracene		1.249	1.214	1.293	1.180	1.337	1.238	5.1
Benzo (g,h,i) perylene		1.239	1.198	1.266	1.188	1.332	1.234	4.7
1,2,4,5-Tetrachlorobenzene		0.626	0.609	0.602	0.557	0.598	0.590	4.3
1,4-Dioxane		0.479	0.499	0.471	0.461	0.509	0.480	3.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1