

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142789.D
 Acq On : 19 Jun 2025 17:38
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Jun 20 04:37:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 04:36:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	80144	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	305306	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	162729	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	277233	20.000	ng	0.00
76) Chrysene-d12	14.051	240	151689	20.000	ng	0.00
86) Perylene-d12	15.545	264	132284	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	49599	10.555	ng	0.00
7) Phenol-d6	6.498	99	60590	10.981	ng	-0.01
23) Nitrobenzene-d5	7.439	82	56291	10.106	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	16892	9.551	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	141565	11.573	ng	0.00
79) Terphenyl-d14	12.992	244	128289	11.771	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.652	88	9589	5.139	ng	95
3) Pyridine	3.434	79	23835	5.102	ng	98
6) Aniline	6.540	93	39088	5.286	ng	97
8) 2-Chlorophenol	6.663	128	27282	5.332	ng	96
10) Phenol	6.516	94	33027	5.369	ng	94
11) bis(2-Chloroethyl)ether	6.610	93	23378	5.122	ng	99
12) 1,3-Dichlorobenzene	6.822	146	30770	5.281	ng	99
13) 1,4-Dichlorobenzene	6.898	146	31311	5.302	ng	99
14) 1,2-Dichlorobenzene	7.051	146	29215	5.175	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	39577	5.503	ng	99
17) 2-Methylphenol	7.128	107	20336	5.199	ng	99
18) Hexachloroethane	7.392	117	10670	5.054	ng	98
19) n-Nitroso-di-n-propyla...	7.281	70	17612	5.061	ng	98
22) Acetophenone	7.281	105	35396	5.240	ng	95
24) Nitrobenzene	7.457	77	26353	5.324	ng	98
25) Isophorone	7.698	82	46673	4.997	ng	99
26) 2-Nitrophenol	7.781	139	12914	4.682	ng	99
27) 2,4-Dimethylphenol	7.810	122	24614	5.242	ng	99
28) bis(2-Chloroethoxy)met...	7.910	93	29467	5.068	ng	99
29) 2,4-Dichlorophenol	8.022	162	22222	5.134	ng	99
30) 1,2,4-Trichlorobenzene	8.104	180	24724	5.171	ng	98
31) Naphthalene	8.187	128	80247	5.344	ng	99
33) 4-Chloroaniline	8.234	127	31626	5.256	ng	98
34) Hexachlorobutadiene	8.304	225	14900	4.918	ng	97
36) 4-Chloro-3-methylphenol	8.710	107	22861	5.113	ng	98
37) 2-Methylnaphthalene	8.875	142	50024	5.281	ng	99
38) 1-Methylnaphthalene	8.975	142	52193	5.307	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	25456	5.419	ng	98
43) 2,4,6-Trichlorophenol	9.157	196	15667	5.154	ng	100
44) 2,4,5-Trichlorophenol	9.192	196	16563	5.045	ng	97
46) 1,1'-Biphenyl	9.339	154	69218	5.478	ng	99
47) 2-Chloronaphthalene	9.363	162	52868	5.670	ng	98
48) 2-Nitroaniline	9.457	65	13076	4.937	ng	96
49) Acenaphthylene	9.781	152	85856	5.477	ng	99
50) Dimethylphthalate	9.633	163	55804	5.170	ng	99
51) 2,6-Dinitrotoluene	9.698	165	12080	5.163	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142789.D
 Acq On : 19 Jun 2025 17:38
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Quant Time: Jun 20 04:37:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 04:36:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
52) Acenaphthene	9.951	154	50726	5.230	ng	97
53) 3-Nitroaniline	9.869	138	12667	4.984	ng	99
55) Dibenzofuran	10.122	168	74998	5.420	ng	99
57) 2,4-Dinitrotoluene	10.104	165	15173	4.904	ng	96
58) Fluorene	10.469	166	59558	5.453	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	13285	4.827	ng	# 97
60) Diethylphthalate	10.333	149	54038	5.040	ng	99
61) 4-Chlorophenyl-phenyle...	10.457	204	29011	5.469	ng	98
62) 4-Nitroaniline	10.475	138	11342	4.971	ng	97
63) Azobenzene	10.616	77	49436	5.276	ng	99
66) n-Nitrosodiphenylamine	10.575	169	51021	5.356	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	15996	4.927	ng	99
68) Hexachlorobenzene	11.016	284	18237	5.059	ng	97
69) Atrazine	11.098	200	12221	4.832	ng	98
71) Phenanthrene	11.433	178	81962	5.522	ng	99
72) Anthracene	11.486	178	84944	5.529	ng	99
73) Carbazole	11.639	167	72067	5.601	ng	100
74) Di-n-butylphthalate	11.969	149	69083	4.850	ng	99
75) Fluoranthene	12.622	202	78721	5.582	ng	99
78) Pyrene	12.851	202	80026	5.739	ng	100
80) Butylbenzylphthalate	13.469	149	18113	4.383	ng	96
81) Benzo(a)anthracene	14.045	228	52785	5.300	ng	100
83) Chrysene	14.080	228	47142	5.084	ng	98
84) Bis(2-ethylhexyl)phtha...	14.027	149	24641	4.002	ng	97
87) Indeno(1,2,3-cd)pyrene	17.045	276	50546	5.168	ng	99
88) Benzo(b)fluoranthene	15.104	252	38493	4.996	ng	99
89) Benzo(k)fluoranthene	15.133	252	39874	5.302	ng	97
90) Benzo(a)pyrene	15.480	252	37779	5.124	ng	99
91) Dibenzo(a,h)anthracene	17.062	278	41301	5.194	ng	100
92) Benzo(g,h,i)perylene	17.504	276	40975	5.165	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 04:37:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 20 04:36:38 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.881 min Scan# 79
 Delta R.T. -0.000 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 80144
 Ion Ratio Lower Upper
 152 100
 150 157.7 127.2 190.8
 115 58.8 46.6 69.8



#2
 1,4-Dioxane
 Concen: 5.139 ng
 RT: 2.652 min Scan# 78
 Delta R.T. 0.012 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38

Tgt Ion: 88 Resp: 9589
 Ion Ratio Lower Upper
 88 100
 58 92.9 69.1 103.7
 43 35.1 27.4 41.2





#3

Pyridine

Concen: 5.102 ng

RT: 3.434 min Scan# 211

Delta R.T. 0.024 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 79 Resp: 23835

Ion Ratio Lower Upper

79 100

52 85.8 69.2 103.8

51 43.7 33.2 49.8



#5

2-Fluorophenol

Concen: 10.555 ng

RT: 5.493 min Scan# 561

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:112 Resp: 49599


Ion Ratio Lower Upper

112 100

64 56.9 48.2 72.2

63 31.1 25.7 38.5









#6
Aniline
Concen: 5.286 ng
RT: 6.540 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 39088
Ion Ratio Lower Upper
93 100
66 41.4 31.8 47.6
65 22.0 16.2 24.2



#7
Phenol-d6
Concen: 10.981 ng
RT: 6.498 min Scan# 732
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion: 99 Resp: 60590
Ion Ratio Lower Upper
99 100
42 19.4 15.9 23.9
71 33.1 27.8 41.8









#8
2-Chlorophenol
Concen: 5.332 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 27282
Ion Ratio Lower Upper
128 100
130 31.3 12.7 52.7
64 43.4 26.5 66.5



#10
Phenol
Concen: 5.369 ng
RT: 6.516 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion: 94 Resp: 33027
Ion Ratio Lower Upper
94 100
65 28.7 11.3 51.3
66 40.5 24.9 64.9









#11
bis(2-Chloroethyl)ether
Concen: 5.122 ng
RT: 6.610 min Scan# 751
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 23378
Ion Ratio Lower Upper
93 100
63 80.6 60.3 100.3
95 34.1 12.5 52.5



#12
1,3-Dichlorobenzene
Concen: 5.281 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:146 Resp: 30770
Ion Ratio Lower Upper
146 100
148 62.9 51.1 76.7
75 27.7 22.7 34.1





#13

1,4-Dichlorobenzene

Concen: 5.302 ng

RT: 6.898 min Scan# 800

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDIC005





Tgt Ion:146 Resp: 31311

Ion Ratio Lower Upper

146 100

148 62.8 50.2 75.2

111 39.8 33.2 49.8



#14

1,2-Dichlorobenzene

Concen: 5.175 ng

RT: 7.051 min Scan# 826

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:146 Resp: 29215


Ion Ratio Lower Upper

146 100

148 63.5 50.4 75.6

111 41.4 34.1 51.1









#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.503 ng
RT: 7.151 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	45	Resp:	39577
Ion Ratio	Lower	Upper	
45	100		
77	12.5	0.0	32.8
79	9.5	0.0	29.8



#17
2-Methylphenol
Concen: 5.199 ng
RT: 7.128 min Scan# 839
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:	107	Resp:	20336
Ion Ratio	Lower	Upper	
107	100		
108	113.6	90.9	136.3
77	47.0	37.3	55.9
79	44.7	37.2	55.8









#18
Hexachloroethane
Concen: 5.054 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDIC005


Tgt Ion:117 Resp: 10670
Ion Ratio Lower Upper
117 100
119 93.8 75.4 113.0
201 97.7 75.9 113.9



#19
n-Nitroso-di-n-propylamine
Concen: 5.061 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion: 70 Resp: 17612
Ion Ratio Lower Upper
70 100
42 66.5 52.1 78.1
101 8.9 8.2 12.2
130 22.0 18.2 27.2









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:136 Resp: 305306
Ion Ratio Lower Upper
136 100
137 10.8 8.4 12.6
54 8.3 6.3 9.5
68 4.8 3.7 5.5



#22
Acetophenone
Concen: 5.240 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:105 Resp: 35396
Ion Ratio Lower Upper
105 100
71 8.5 6.5 9.7
51 41.9 29.8 44.8
120 23.4 18.2 27.2





#23

Nitrobenzene-d5

Concen: 10.106 ng

RT: 7.439 min Scan# 892

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 82 Resp: 56291

Ion Ratio Lower Upper

82 100

128 48.9 37.7 56.5

54 56.6 45.7 68.5



#24

Nitrobenzene

Concen: 5.324 ng

RT: 7.457 min Scan# 895

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion: 77 Resp: 26353


Ion Ratio Lower Upper

77 100

123 49.9 39.2 58.8

65 13.3 11.6 17.4









#25
Isophorone
Concen: 4.997 ng
RT: 7.698 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 46673
Ion Ratio Lower Upper
82 100
95 8.1 6.2 9.2
138 20.8 16.2 24.4



#26
2-Nitrophenol
Concen: 4.682 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:139 Resp: 12914
Ion Ratio Lower Upper
139 100
109 27.7 22.7 34.1
65 48.8 38.6 58.0









#27
2,4-Dimethylphenol
Concen: 5.242 ng
RT: 7.810 min Scan# 910
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:122 Resp: 24614
Ion Ratio Lower Upper
122 100
107 110.7 88.0 132.0
121 56.1 46.3 69.5



#28
bis(2-Chloroethoxy)methane
Concen: 5.068 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion: 93 Resp: 29467
Ion Ratio Lower Upper
93 100
95 32.2 26.2 39.2
123 12.1 9.9 14.9





#29

2,4-Dichlorophenol

Concen: 5.134 ng

RT: 8.022 min Scan# 991

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 22222

Ion Ratio Lower Upper

162 100

164 63.2 44.0 84.0

98 34.4 15.4 55.4



#30

1,2,4-Trichlorobenzene

Concen: 5.171 ng

RT: 8.104 min Scan# 1005

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:180 Resp: 24724


Ion Ratio Lower Upper

180 100

182 97.8 76.4 114.6

145 31.0 25.8 38.8









#31
Naphthalene
Concen: 5.344 ng
RT: 8.187 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 80247
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 12.4 10.4 15.6



#33
4-Chloroaniline
Concen: 5.256 ng
RT: 8.234 min Scan# 1027
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:127 Resp: 31626
Ion Ratio Lower Upper
127 100
129 31.2 25.9 38.9
65 29.9 24.6 36.8
92 18.5 15.4 23.0









#34
Hexachlorobutadiene
Concen: 4.918 ng
RT: 8.304 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:225 Resp: 14900
Ion Ratio Lower Upper
225 100
223 65.5 49.9 74.9
227 65.7 51.4 77.0



#36
4-Chloro-3-methylphenol
Concen: 5.113 ng
RT: 8.710 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:107 Resp: 22861
Ion Ratio Lower Upper
107 100
144 26.1 21.3 31.9
142 80.7 66.6 100.0









#37
2-Methylnaphthalene
Concen: 5.281 ng
RT: 8.875 min Scan# 1136
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:142 Resp: 50024
Ion Ratio Lower Upper
142 100
141 87.6 70.9 106.3



#38
1-Methylnaphthalene
Concen: 5.307 ng
RT: 8.975 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:142 Resp: 52193
Ion Ratio Lower Upper
142 100
141 91.2 73.0 109.4









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:164 Resp: 162729
Ion Ratio Lower Upper
164 100
162 101.1 81.8 122.8
160 44.2 36.0 54.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.419 ng
RT: 9.045 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:216 Resp: 25456
Ion Ratio Lower Upper
216 100
214 78.0 63.7 95.5
179 21.1 17.6 26.4
108 16.2 14.3 21.5









#42
2,4,6-Tribromophenol
Concen: 9.551 ng
RT: 10.710 min Scan# 1449
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:330 Resp: 16892
Ion Ratio Lower Upper
330 100
332 96.5 75.0 112.6
141 34.5 25.3 37.9



#43
2,4,6-Trichlorophenol
Concen: 5.154 ng
RT: 9.157 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:196 Resp: 15667
Ion Ratio Lower Upper
196 100
198 95.2 76.2 114.2
200 30.2 24.1 36.1









#44
2,4,5-Trichlorophenol
Concen: 5.045 ng
RT: 9.192 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:196 Resp: 16563
Ion Ratio Lower Upper
196 100
198 94.8 76.6 115.0
97 51.1 38.2 57.2
132 32.2 23.4 35.0



#45
2-Fluorobiphenyl
Concen: 11.573 ng
RT: 9.239 min Scan# 1198
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:172 Resp: 141565
Ion Ratio Lower Upper
172 100
171 34.8 28.2 42.4
170 22.5 18.5 27.7









#46
 1,1'-Biphenyl
 Concen: 5.478 ng
 RT: 9.339 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:154 Resp: 69218
 Ion Ratio Lower Upper
 154 100
 153 39.6 20.3 60.3
 76 9.5 0.0 30.3



#47
 2-Chloronaphthalene
 Concen: 5.670 ng
 RT: 9.363 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38

Tgt Ion:162 Resp: 52868
 Ion Ratio Lower Upper
 162 100
 127 38.4 32.2 48.2
 164 32.4 26.1 39.1









#48
2-Nitroaniline
Concen: 4.937 ng
RT: 9.457 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	65	Resp:	13076
Ion	Ratio	Lower	Upper
65	100		
92	69.4	53.8	80.6
138	106.9	89.6	134.4



#49
Acenaphthylene
Concen: 5.477 ng
RT: 9.781 min Scan# 1290
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38


Tgt Ion:	152	Resp:	85856
Ion	Ratio	Lower	Upper
152	100		
151	19.7	16.1	24.1
153	12.8	10.4	15.6








#50
Dimethylphthalate
Concen: 5.170 ng
RT: 9.633 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38


Instrument :
BNA_F
ClientSampleId :
SSTDICC005





Tgt Ion:163 Resp: 55804
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.2 8.0 12.0




#51
2,6-Dinitrotoluene
Concen: 5.163 ng
RT: 9.698 min Scan# 1276
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38



Tgt Ion:165 Resp: 12080
Ion Ratio Lower Upper
165 100
63 56.1 44.2 66.4
89 52.1 39.8 59.8









#52
Acenaphthene
Concen: 5.230 ng
RT: 9.951 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:154 Resp: 50726
Ion Ratio Lower Upper
154 100
153 115.3 90.0 135.0
152 54.3 41.9 62.9



#53
3-Nitroaniline
Concen: 4.984 ng
RT: 9.869 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:138 Resp: 12667
Ion Ratio Lower Upper
138 100
108 10.7 8.1 12.1
92 108.0 85.1 127.7









#55
Dibenzenofuran
Concen: 5.420 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:168 Resp: 74998
Ion Ratio Lower Upper
168 100
139 38.5 30.2 45.4
169 13.8 10.8 16.2



#57
2,4-Dinitrotoluene
Concen: 4.904 ng
RT: 10.104 min Scan# 1345
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:165 Resp: 15173
Ion Ratio Lower Upper
165 100
63 45.7 33.9 50.9
89 65.9 50.6 75.8
182 2.7 2.3 3.5





#58

Fluorene

Concen: 5.453 ng

RT: 10.469 min Scan# 1408

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDIC005





Tgt Ion:166 Resp: 59558

Ion Ratio Lower Upper

166 100

165 94.0 75.2 112.8

167 13.0 10.6 15.8



#59

2,3,4,6-Tetrachlorophenol

Concen: 4.827 ng

RT: 10.245 min Scan# 1369

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Tgt Ion:232 Resp: 13285




Ion Ratio Lower Upper


232 100

131 44.7 37.6 56.4

130 1.7 1.7 2.5#

166 30.3 25.8 38.6









#60
Diethylphthalate
Concen: 5.040 ng
RT: 10.333 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	149	Resp:	54038
Ion Ratio	Lower	Upper	
149	100		
177	21.9	17.0	25.6
150	12.6	9.8	14.6



#61
4-Chlorophenyl-phenylether
Concen: 5.469 ng
RT: 10.457 min Scan# 1405
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:	204	Resp:	29011
Ion Ratio	Lower	Upper	
204	100		
206	32.2	26.3	39.5
141	59.2	49.1	73.7





#62

4-Nitroaniline

Concen: 4.971 ng

RT: 10.475 min Scan# 1408

Delta R.T. -0.012 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:138 Resp: 11342

Ion Ratio Lower Upper

138 100

92 45.5 28.4 68.4

108 71.3 50.2 90.2



#63

Azobenzene

Concen: 5.276 ng

RT: 10.616 min Scan# 1432

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Tgt Ion: 77 Resp: 49436




Ion Ratio Lower Upper


77 100

182 26.8 7.7 47.7

105 21.4 1.9 41.9

51 34.2 13.3 53.3









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:188 Resp: 277233
Ion Ratio Lower Upper
188 100
94 8.2 6.7 10.1
80 8.3 6.9 10.3



#66
n-Nitrosodiphenylamine
Concen: 5.356 ng
RT: 10.575 min Scan# 1425
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:169 Resp: 51021
Ion Ratio Lower Upper
169 100
168 65.7 51.9 77.9
167 35.1 28.5 42.7









#67
4-Bromophenyl-phenylether
Concen: 4.927 ng
RT: 10.951 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:248 Resp: 15996
Ion Ratio Lower Upper
248 100
250 100.4 79.8 119.8
141 73.1 57.8 86.6



#68
Hexachlorobenzene
Concen: 5.059 ng
RT: 11.016 min Scan# 1500
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:284 Resp: 18237
Ion Ratio Lower Upper
284 100
142 34.6 25.5 38.3
249 31.2 24.9 37.3









#69
Atrazine
Concen: 4.832 ng
RT: 11.098 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:200 Resp: 12221
Ion Ratio Lower Upper
200 100
173 27.6 7.4 47.4
215 46.2 28.4 68.4



#71
Phenanthrene
Concen: 5.522 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:178 Resp: 81962
Ion Ratio Lower Upper
178 100
176 18.4 15.1 22.7
179 14.8 12.0 18.0





#72

Anthracene

Concen: 5.529 ng

RT: 11.486 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:178 Resp: 84944

Ion Ratio Lower Upper

178 100

176 17.6 14.6 21.8

179 15.1 12.4 18.6



#73

Carbazole

Concen: 5.601 ng

RT: 11.639 min Scan# 1606

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:167 Resp: 72067


Ion Ratio Lower Upper

167 100

166 21.1 17.0 25.4

139 12.5 10.2 15.2





#74

Di-n-butylphthalate

Concen: 4.850 ng

RT: 11.969 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:149 Resp: 69083

Ion Ratio Lower Upper

149 100

150 8.7 7.4 11.0

104 4.2 3.5 5.3



#75

Fluoranthene

Concen: 5.582 ng

RT: 12.622 min Scan# 1773

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:202 Resp: 78721


Ion Ratio Lower Upper

202 100

101 10.0 0.0 29.7

203 17.2 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.051 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:240 Resp: 151689

Ion Ratio Lower Upper

240 100

120 11.5 8.2 12.2

236 24.9 19.8 29.8



#78

Pyrene

Concen: 5.739 ng

RT: 12.851 min Scan# 1812

Delta R.T. -0.006 min

Lab File: BF142789.D

Acq: 19 Jun 2025 17:38




Tgt Ion:202 Resp: 80026


Ion Ratio Lower Upper

202 100

200 19.2 15.7 23.5


203 17.2 13.8 20.6








#79
 Terphenyl-d14
 Concen: 11.771 ng
 RT: 12.992 min Scan# 1836
 Delta R.T. -0.006 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005







Tgt Ion:244 Resp: 128289
 Ion Ratio Lower Upper
 244 100
 212 6.7 5.4 8.0
 122 10.2 8.2 12.2



#80
 Butylbenzylphthalate
 Concen: 4.383 ng
 RT: 13.469 min Scan# 1917
 Delta R.T. 0.000 min
 Lab File: BF142789.D
 Acq: 19 Jun 2025 17:38

Tgt Ion:149 Resp: 18113
 Ion Ratio Lower Upper
 149 100
 91 62.1 52.9 79.3
 206 19.5 16.3 24.5









#81
Benzo(a)anthracene
Concen: 5.300 ng
RT: 14.045 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:228 Resp: 52785
Ion Ratio Lower Upper
228 100
226 26.5 21.2 31.8
229 19.9 15.6 23.4



#83
Chrysene
Concen: 5.084 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:228 Resp: 47142
Ion Ratio Lower Upper
228 100
226 28.7 23.7 35.5
229 18.9 15.7 23.5









#84
Bis(2-ethylhexyl)phthalate
Concen: 4.002 ng
RT: 14.027 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 24641
Ion Ratio Lower Upper
149 100
167 28.3 21.4 32.0
279 4.0 3.0 4.6



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.545 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:264 Resp: 132284
Ion Ratio Lower Upper
264 100
260 22.7 18.1 27.1
265 21.3 16.6 25.0









#87
Indeno(1,2,3-cd)pyrene
Concen: 5.168 ng
RT: 17.045 min Scan# 2195
Delta R.T. -0.018 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:276 Resp: 50546
Ion Ratio Lower Upper
276 100
138 25.7 20.2 30.4
277 24.9 19.8 29.6



#88
Benzo(b)fluoranthene
Concen: 4.996 ng
RT: 15.104 min Scan# 2195
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:252 Resp: 38493
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 9.7 7.4 11.0









#89
Benzo(k)fluoranthene
Concen: 5.302 ng
RT: 15.133 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 39874
Ion Ratio Lower Upper
252 100
253 23.3 17.2 25.8
125 9.4 7.4 11.2



#90
Benzo(a)pyrene
Concen: 5.124 ng
RT: 15.480 min Scan# 2259
Delta R.T. -0.006 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:252 Resp: 37779
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 10.3 8.6 12.8









#91
Dibenzo(a,h)anthracene
Concen: 5.194 ng
RT: 17.062 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:278 Resp: 41301
Ion Ratio Lower Upper
278 100
139 17.1 13.5 20.3
279 23.6 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 5.165 ng
RT: 17.504 min Scan# 2603
Delta R.T. -0.012 min
Lab File: BF142789.D
Acq: 19 Jun 2025 17:38

Tgt Ion:276 Resp: 40975
Ion Ratio Lower Upper
276 100
277 23.7 18.8 28.2
138 22.6 17.6 26.4

