

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061925

Review By	Rahul	Review On	6/23/2025 3:55:52 PM
Supervise By	Jagrut	Supervise On	6/24/2025 3:17:49 PM
SubDirectory	BF061925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12670,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142787.D	19 Jun 2025 16:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142788.D	19 Jun 2025 17:07		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142789.D	19 Jun 2025 17:38		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF142790.D	19 Jun 2025 18:08		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142791.D	19 Jun 2025 18:39		RC/JU	Ok
6	SSTDICC040	SSTDICC040	BF142792.D	19 Jun 2025 19:09	Calibration is Good for 8270 E, 8270 DOD and 625.1 methods.	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142793.D	19 Jun 2025 19:40		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142794.D	19 Jun 2025 20:10		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142795.D	19 Jun 2025 20:40	Compound #09 removed from 80 PPM.	RC/JU	Ok
10	SSTDICV040	ICVBF061925	BF142796.D	19 Jun 2025 21:10		RC/JU	Ok
11	PB168536BL	PB168536BL	BF142797.D	19 Jun 2025 22:10		RC/JU	Ok
12	PB168536BS	PB168536BS	BF142798.D	19 Jun 2025 22:40		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062425

Review By	Rahul	Review On	6/25/2025 10:47:43 AM
Supervise By	Jagrut	Supervise On	6/25/2025 11:19:22 AM
SubDirectory	BF062425	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6836 S12671,10ul/1000ul sample SP6770

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142810.D	24 Jun 2025 09:46		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142811.D	24 Jun 2025 10:16		RC/JU	Ok
3	PB168582BL	PB168582BL	BF142812.D	24 Jun 2025 10:46		RC/JU	Ok
4	PB168582BS	PB168582BS	BF142813.D	24 Jun 2025 11:17		RC/JU	Ok,M
5	Q2388-01	TP-12	BF142814.D	24 Jun 2025 11:51		RC/JU	Ok
6	Q2380-01	72-11948	BF142815.D	24 Jun 2025 12:22	Internal Standard Fail	RC/JU	Ok,M
7	Q2380-03	VNJ-211	BF142816.D	24 Jun 2025 12:53		RC/JU	Ok,M
8	Q2381-01	291431	BF142817.D	24 Jun 2025 13:24		RC/JU	Ok,M
9	Q2381-05	72-11929	BF142818.D	24 Jun 2025 13:54	Internal Standard Fail	RC/JU	Ok,M
10	Q2385-01	A5311	BF142819.D	24 Jun 2025 14:57		RC/JU	Ok
11	Q2372-01	GAV1W	BF142820.D	24 Jun 2025 15:28		RC/JU	Ok
12	Q2386-01	SU-1-062025	BF142821.D	24 Jun 2025 15:59	Internal Standard Fail, Run 10X Dilution and then decide further dilution	RC/JU	Dilution
13	Q2392-01	S-1	BF142822.D	24 Jun 2025 16:30		RC/JU	Ok
14	Q2393-11	PARK AVE -6	BF142823.D	24 Jun 2025 17:01		RC/JU	Ok
15	Q2393-05	PARK AVE -3	BF142824.D	24 Jun 2025 17:33		RC/JU	Ok,M
16	Q2393-07	PARK AVE -4	BF142825.D	24 Jun 2025 18:03		RC/JU	Ok,M
17	Q2393-09	PARK AVE -5	BF142826.D	24 Jun 2025 18:34		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062425

Review By	Rahul	Review On	6/25/2025 10:47:43 AM		
Supervise By	Jagrut	Supervise On	6/25/2025 11:19:22 AM		
SubDirectory	BF062425	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840			
CCC		SP6836			
Internal Standard/PEM		S12671,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2398-06	ARS20-0036	BF142827.D	24 Jun 2025 19:05	Internal Standard Fail	RC/JU	Ok,M
19	Q2398-03	M00-25-0179	BF142828.D	24 Jun 2025 19:36	Internal Standard Fail	RC/JU	Ok,M
20	Q2381-03	VNJ-207	BF142829.D	24 Jun 2025 20:07		RC/JU	Ok,M
21	Q2380-05	RBR-200059	BF142830.D	24 Jun 2025 20:38		RC/JU	Ok,M
22	Q2384-01	OK-01-6202025	BF142831.D	24 Jun 2025 21:08		RC/JU	Ok,M
23	Q2383-01	RT-5376	BF142832.D	24 Jun 2025 21:38	Internal Standard Fail	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062525

Review By	Rahul	Review On	6/26/2025 9:14:40 AM
Supervise By	Jagrut	Supervise On	6/26/2025 10:48:09 AM
SubDirectory	BF062525	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12671,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142833.D	25 Jun 2025 12:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142834.D	25 Jun 2025 13:25		RC/JU	Ok,M
3	PB168592BL	PB168592BL	BF142835.D	25 Jun 2025 13:55		RC/JU	Ok
4	PB168592BS	PB168592BS	BF142836.D	25 Jun 2025 14:25		RC/JU	Ok,M
5	PB168592BSD	PB168592BSD	BF142837.D	25 Jun 2025 14:55		RC/JU	Ok,M
6	PB168601BL	PB168601BL	BF142838.D	25 Jun 2025 15:26		RC/JU	Ok
7	PB168601BS	PB168601BS	BF142839.D	25 Jun 2025 15:57		RC/JU	Ok
8	Q2393-01	PARK AVE -1	BF142840.D	25 Jun 2025 16:32		RC/JU	Ok
9	Q2393-03	PARK AVE -2	BF142841.D	25 Jun 2025 17:02		RC/JU	Ok
10	Q2394-01	MH-K/L	BF142842.D	25 Jun 2025 17:33		RC/JU	Ok
11	Q2394-01MS	MH-K/LMS	BF142843.D	25 Jun 2025 18:04		RC/JU	Ok
12	Q2394-01MSD	MH-K/LMSD	BF142844.D	25 Jun 2025 18:35		RC/JU	Ok
13	Q2386-02	SU-1-062025	BF142845.D	25 Jun 2025 19:06		RC/JU	Ok
14	Q2386-02MS	SU-1-062025MS	BF142846.D	25 Jun 2025 19:37		RC/JU	Ok
15	Q2386-02MSD	SU-1-062025MSD	BF142847.D	25 Jun 2025 20:07		RC/JU	Ok
16	Q2388-04	TP-12	BF142848.D	25 Jun 2025 20:38		RC/JU	Ok
17	Q2389-04	MH-J-I	BF142849.D	25 Jun 2025 21:09		RC/JU	Ok
18	Q2399-01	TP-13	BF142850.D	25 Jun 2025 21:40		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062525

Review By	Rahul	Review On	6/26/2025 9:14:40 AM		
Supervise By	Jagrut	Supervise On	6/26/2025 10:48:09 AM		
SubDirectory	BF062525	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12671,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2399-05	EP-7	BF142851.D	25 Jun 2025 22:10		RC/JU	Ok
20	Q2386-01DL	SU-1-062025DL	BF142852.D	25 Jun 2025 22:41	Need Further 5X Dilution	RC/JU	Dilution
21	Q2386-01DL2	SU-1-062025DL2	BF142853.D	25 Jun 2025 23:11		RC/JU	Ok,M
22	Q2412-01	TR-05-062425	BF142854.D	25 Jun 2025 23:41		RC/JU	Ok,M
23	Q2403-03	LAW-25-0093	BF142855.D	26 Jun 2025 00:12	Surrogate Failed.	RC/JU	Ok

M : Manual Integration