

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG NO.: Q2372

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/24/2025

Lab File ID: BF142811.D Time Analyzed: 10:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	88364	6.88	335549	8.16	179366	9.92
UPPER LIMIT	176728	7.38	671098	8.663	358732	10.421
LOWER LIMIT	44182	6.38	167775	7.663	89683	9.421
EPA SAMPLE NO.						
01 GAV1W	69183	6.88	242235	8.16	106484	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.