

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG NO.: Q2372

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/24/2025

Lab File ID: BF142811.D Time Analyzed: 10:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	300319	11.41	163667	14.056	187854	15.545
UPPER LIMIT	600638	11.91	327334	14.556	375708	16.045
LOWER LIMIT	150160	10.91	81833.5	13.556	93927	15.045
EPA SAMPLE NO.						
GAV1W	179849	11.41	193130	14.06	144569	15.55

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.