

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2372 SAS No.: Q2372 SDG NO.: Q2372
Lab File ID: VX046758.D Date Analyzed: 06/19/2025
Instrument ID: MSVOA_X Time Analyzed: 09:43
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137182	5.56	224265	6.76	197740	10.05
UPPER LIMIT	274364	6.056	448530	7.263	395480	10.549
LOWER LIMIT	68591	5.056	112133	6.263	98870	9.549
EPA SAMPLE NO.						
GAV1W	139696	5.56	239120	6.77	216781	10.06
GAV2W	168816	5.56	291352	6.77	260859	10.05
VX0619WBL01	124937	5.56	212303	6.77	191555	10.06
VX0619WBS01	137936	5.56	227140	6.77	207463	10.06
VX0619WBSD01	120371	5.56	199496	6.76	183807	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.