

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062125\
 Data File : BN037365.D
 Acq On : 21 Jun 2025 01:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 21 01:41:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 23:41:54 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.407	0.399	2.0	101	0.00
3	n-Nitrosodimethylamine	0.373	0.370	0.8	93	0.00
4 S	2-Fluorophenol	0.799	0.740	7.4	98	0.00
5 S	Phenol-d6	0.823	0.753	8.5	97	0.00
6	bis(2-Chloroethyl)ether	0.730	0.673	7.8	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.323	0.318	1.5	100	0.00
9	Naphthalene	1.056	1.038	1.7	99	0.00
10	Hexachlorobutadiene	0.417	0.452	-8.4	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.648	0.670	-3.4	100	0.00
12	2-Methylnaphthalene	0.736	0.709	3.7	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.235	0.218	7.2	94	0.00
15 S	2-Fluorobiphenyl	1.757	1.777	-1.1	99	0.00
16	Acenaphthylene	1.682	1.615	4.0	100	0.00
17	Acenaphthene	1.107	1.059	4.3	99	0.00
18	Fluorene	1.556	1.496	3.9	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.096	0.084	12.5	105	0.01
21	4-Bromophenyl-phenylether	0.285	0.282	1.1	101	0.01
22	Hexachlorobenzene	0.310	0.316	-1.9	100	0.00
23	Atrazine	0.227	0.207	8.8	94	0.00
24	Pentachlorophenol	0.154	0.140	9.1	102	0.00
25	Phenanthrene	1.158	1.093	5.6	99	0.00
26	Anthracene	1.068	0.961	10.0	97	0.00
27 SURR	Fluoranthene-d10	1.148	1.189	-3.6	102	0.00
28	Fluoranthene	1.464	1.410	3.7	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.625	1.577	3.0	101	0.00
31 S	Terphenyl-d14	0.912	0.903	1.0	104	0.00
32	Benzo(a)anthracene	1.315	1.237	5.9	109	0.00
33	Chrysene	1.594	1.559	2.2	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.540	0.520	3.7	103	0.00
35 I	Perylene-d12	1.000	1.000	0.0	115	0.00
36	Indeno(1,2,3-cd)pyrene	1.794	1.712	4.6	116	0.00
37	Benzo(b)fluoranthene	1.472	1.336	9.2	106	0.00
38	Benzo(k)fluoranthene	1.597	1.587	0.6	116	0.00
39 C	Benzo(a)pyrene	1.319	1.253	5.0	113	0.00
40	Dibenzo(a,h)anthracene	1.348	1.307	3.0	122	0.00
41	Benzo(g,h,i)perylene	1.603	1.559	2.7	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4