

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2373 SAS No.: Q2373 SDG NO.: Q2373

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/26/2025

Lab File ID: BN037403.D Time Analyzed: 21:38

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1253	7.56	2741	10.34	1842	14.21
UPPER LIMIT	2506	8.06	5482	10.84	3684	14.713
LOWER LIMIT	626.5	7.06	1370.5	9.84	921	13.713
EPA SAMPLE NO.						
01 RW5-SP201-20250619	2329	7.57	5147	10.34	3323	14.21
02 RW5-SP303-20250619	2127	7.57	4523	10.34	3031	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.