

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2373 SAS No.: Q2373 SDG NO.: Q2373

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/20/2025

Lab File ID: BN037355.D Time Analyzed: 18:03

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5776	16.971	4813	21.171	4943	23.354
UPPER LIMIT	11552	17.471	9626	21.671	9886	23.854
LOWER LIMIT	2888	16.471	2406.5	20.671	2471.5	22.854
EPA SAMPLE NO.						
01 PB168563BL	4864	16.98	4288	21.17	3457	23.36
02 PB168563BS	4830	16.97	3875	21.17	2749	23.35
03 PB168563BSD	5035	16.97	4193	21.16	4470	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.