

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2373 SDG NO.: Q2373

Lab File ID: BN037416.D

DFTPP Injection Date: 06/27/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1.1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	79.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037417.D	06/27/2025	10:09
RW5-SP100-20250619	Q2373-01	BN037420.D	06/27/2025	11:57
RW5-SP100-20250619DL	Q2373-01DL	BN037421.D	06/27/2025	12:34
SSTDCCC0.4EC	SSTDCCC0.4	BN037423.D	06/27/2025	15:34