

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2374

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN037363.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168563BS	1,4-Dioxane	0.4	0.32	ug/L	80				70	130	