

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2375

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136237

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136237BS							
TDS	mg/L	100	95.0		95	1	90-110	06/23/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2375

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136238

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136238BS							
TSS	mg/L	550	534		97	1	90-110	06/24/2025