

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3840	4000	96	90 - 110	P	06/24/2025	18:43	LB136255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4870	5000	98	90 - 110	P	06/24/2025	19:34	LB136255
CCV02	Iron	4880	5000	98	90 - 110	P	06/24/2025	20:37	LB136255
CCV03	Iron	4950	5000	99	90 - 110	P	06/24/2025	21:29	LB136255
CCV04	Iron	4860	5000	97	90 - 110	P	06/24/2025	22:22	LB136255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLCCV01	Iron	105	100	105	80 - 120	P	06/24/2025	22:47	LB136255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Iron	4750	5000	95	90 - 110	P	06/24/2025	23:46	LB136255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	90 - 110	P	06/25/2025	13:08	LB136273

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	92.0	100	92	80 - 120	P	06/25/2025	13:20	LB136273

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5120	5000	102	90 - 110	P	06/25/2025	16:04	LB136273
CCV02	Iron	5120	5000	102	90 - 110	P	06/25/2025	16:57	LB136273
CCV03	Iron	5130	5000	103	90 - 110	P	06/25/2025	18:00	LB136273
CCV04	Iron	5150	5000	103	90 - 110	P	06/25/2025	18:55	LB136273
CCV05	Iron	5120	5000	102	90 - 110	P	06/25/2025	19:52	LB136273
CCV06	Iron	5040	5000	101	90 - 110	P	06/25/2025	20:46	LB136273
CCV07	Iron	5050	5000	101	90 - 110	P	06/25/2025	21:12	LB136273
