



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	102	100	102	65 - 135	P	06/24/2025	19:12	LB136255
CRI01	Iron	107	100	107	65 - 135	P	06/25/2025	15:24	LB136273