

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2375 **SAS No.:** Q2375
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	100000	100000	100	85600	116500	06/24/2025	19:17	LB136255
ICSAB01	Iron	99000	99000	100	84400	114500	06/24/2025	19:21	LB136255
ICSA01	Iron	104000	100000	104	85600	116500	06/25/2025	15:29	LB136273
ICSAB01	Iron	101000	99000	102	84400	114500	06/25/2025	15:51	LB136273