

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2375 SAS No.: Q2375 SDG No.: Q2375

Instrument ID: BNA_N Calibration Date/Time: 06/21/2025 01:17

Lab File ID: BN037365.D Init. Calib. Date(s): 06/20/2025 06/20/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 16:51 20:27

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.648	0.670		3.4	50.0
Fluoranthene-d10	1.148	1.189		3.6	50.0
2-Fluorophenol	0.799	0.740		-7.4	50.0
Phenol-d6	0.823	0.753		-8.5	50.0
Nitrobenzene-d5	0.323	0.318		-1.5	50.0
2-Fluorobiphenyl	1.757	1.777		1.1	50.0
2,4,6-Tribromophenol	0.235	0.218		-7.2	50.0
Terphenyl-d14	0.912	0.903		-1.0	50.0
1,4-Dioxane	0.407	0.399		-2.0	50.0

All other compounds must meet a minimum RRF of 0.010.