

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062125\
 Data File : BN037365.D
 Acq On : 21 Jun 2025 01:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/23/2025
 Supervised By :Jagrut Upadhyay 06/24/2025

Quant Time: Jun 21 01:41:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 23:41:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.568	152	1894	0.400	ng	0.00
7) Naphthalene-d8	10.340	136	4150	0.400	ng	# 0.00
13) Acenaphthene-d10	14.213	164	2786	0.400	ng	0.00
19) Phenanthrene-d10	16.971	188	5748	0.400	ng	# 0.00
29) Chrysene-d12	21.162	240	5140	0.400	ng	0.00
35) Perylene-d12	23.351	264	5682	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.163	112	1402	0.370	ng	0.00
5) Phenol-d6	6.752	99	1426	0.366	ng	0.00
8) Nitrobenzene-d5	8.717	82	1321	0.394	ng	0.00
11) 2-Methylnaphthalene-d10	11.945	152	2780	0.413	ng	0.00
14) 2,4,6-Tribromophenol	15.718	330	608	0.372	ng	0.00
15) 2-Fluorobiphenyl	12.838	172	4951	0.405	ng	0.00
27) Fluoranthene-d10	19.008	212	6836	0.414	ng	0.00
31) Terphenyl-d14	19.616	244	4639	0.396	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.104	88	755	0.392	ng	95
3) n-Nitrosodimethylamine	3.422	42	701	0.397	ng	# 78
6) bis(2-Chloroethyl)ether	7.004	93	1274	0.368	ng	100
9) Naphthalene	10.394	128	4308	0.393	ng	# 90
10) Hexachlorobutadiene	10.682	225	1876	0.434	ng	# 100
12) 2-Methylnaphthalene	12.021	142	2941	0.385	ng	93
16) Acenaphthylene	13.935	152	4498	0.384	ng	98
17) Acenaphthene	14.277	154	2949	0.383	ng	98
18) Fluorene	15.272	166	4167	0.385	ng	99
20) 4,6-Dinitro-2-methylph...	15.379	198	483	0.350	ng	85
21) 4-Bromophenyl-phenylether	16.177	248	1621	0.396	ng	# 94
22) Hexachlorobenzene	16.276	284	1819	0.408	ng	99
23) Atrazine	16.450	200	1187	0.364	ng	# 87
24) Pentachlorophenol	16.624	266	805	0.363	ng	96
25) Phenanthrene	17.009	178	6282	0.377	ng	99
26) Anthracene	17.108	178	5522	0.360	ng	99
28) Fluoranthene	19.036	202	8102	0.385	ng	98
30) Pyrene	19.403	202	8105	0.388	ng	100
32) Benzo(a)anthracene	21.153	228	6359	0.376	ng	98
33) Chrysene	21.198	228	8015	0.391	ng	98
34) Bis(2-ethylhexyl)phtha...	21.090	149	2672	0.385	ng	# 98
36) Indeno(1,2,3-cd)pyrene	25.535	276	9725	0.382	ng	# 92
37) Benzo(b)fluoranthene	22.696	252	7592m	0.363	ng	
38) Benzo(k)fluoranthene	22.740	252	9018	0.398	ng	94
39) Benzo(a)pyrene	23.255	252	7122	0.380	ng	# 90
40) Dibenzo(a,h)anthracene	25.553	278	7424	0.388	ng	# 85
41) Benzo(g,h,i)perylene	26.199	276	8859	0.389	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062125\
 Data File : BN037365.D
 Acq On : 21 Jun 2025 01:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/23/2025
 Supervised By :Jagrut Upadhyay 06/24/2025


Quant Time: Jun 21 01:41:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 23:41:54 2025
 Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.568 min Scan# 649
 Delta R.T. -0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC




Tgt Ion:152 Resp: 1894
 Ion Ratio Lower Upper
 152 100
 150 149.8 112.7 169.1
 115 55.5 45.9 68.9

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 06/23/2025
 Supervised By :Jagrut Upadhyay 06/24/2025




#2
 1,4-Dioxane
 Concen: 0.392 ng
 RT: 3.104 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17



Tgt Ion: 88 Resp: 755
 Ion Ratio Lower Upper
 88 100
 43 22.8 21.0 31.6
 58 49.9 38.0 57.0





#3

n-Nitrosodimethylamine

Concen: 0.397 ng

RT: 3.422 min Scan# 75

Delta R.T. 0.000 min

Lab File: BN037365.D


Acq: 21 Jun 2025 01:17

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion: 42 Resp: 70




Ion	Ratio	Lower	Upper
42	100		
74	129.2	84.3	126.5
44	9.3	5.0	7.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/23/2025

Supervised By :Jagrut Upadhyay 06/24/2025



#4

2-Fluorophenol

Concen: 0.370 ng

RT: 5.163 min Scan# 316




Delta R.T. 0.000 min


Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Tgt Ion:112 Resp: 1402


Ion	Ratio	Lower	Upper
112	100		
64	52.2	38.7	58.1
63	31.0	26.4	39.6





#5
Phenol-d6
Concen: 0.366 ng
RT: 6.752 min Scan# 536
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 99 Resp: 1420
Ion Ratio Lower Upper
99 100
42 18.2 19.8 29.8
71 44.0 42.6 64.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.368 ng
RT: 7.004 min Scan# 571
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion: 93 Resp: 1274
Ion Ratio Lower Upper
93 100
63 66.6 53.2 79.8
95 33.8 27.3 40.9





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.340 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:136 Resp: 4150
Ion Ratio Lower Upper
136 100
137 12.6 12.2 18.2
54 8.3 8.8 13.2
68 6.8 7.0 10.4#




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#8
Nitrobenzene-d5
Concen: 0.394 ng
RT: 8.717 min Scan# 790
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion: 82 Resp: 1321
Ion Ratio Lower Upper
82 100
128 41.4 42.5 63.7#
54 44.6 43.2 64.8





#9
Naphthalene
Concen: 0.393 ng
RT: 10.394 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:128 Resp: 4308
Ion Ratio Lower Upper
128 100
129 12.7 14.0 21.0
127 15.2 15.8 23.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025




#10
Hexachlorobutadiene
Concen: 0.434 ng
RT: 10.682 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Tgt Ion:225 Resp: 1876
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.9 50.3 75.5





#11
2-Methylnaphthalene-d10
Concen: 0.413 ng
RT: 11.945 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:152 Resp: 2780
Ion Ratio Lower Upper
152 100
151 21.7 17.4 26.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025




#12
2-Methylnaphthalene
Concen: 0.385 ng
RT: 12.021 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Tgt Ion:142 Resp: 2941
Ion Ratio Lower Upper
142 100
141 90.9 70.2 105.2
115 43.9 43.0 64.4





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.213 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 2780
Ion Ratio Lower Upper
164 100
162 105.6 81.5 122.3
160 53.2 43.0 64.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025




#14
2,4,6-Tribromophenol
Concen: 0.372 ng
RT: 15.718 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Tgt Ion:330 Resp: 608
Ion Ratio Lower Upper
330 100
332 102.3 78.4 117.6
141 35.9 24.4 36.6





#15
2-Fluorobiphenyl
Concen: 0.405 ng
RT: 12.838 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:172 Resp: 495
Ion Ratio Lower Upper
172 100
171 35.8 30.8 46.2
170 25.4 21.9 32.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#16
Acenaphthylene
Concen: 0.384 ng
RT: 13.935 min Scan# 1434
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion:152 Resp: 4498
Ion Ratio Lower Upper
152 100
151 19.8 16.6 24.8
153 13.1 10.2 15.2





#17
Acenaphthene
Concen: 0.383 ng
RT: 14.277 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 2949
Ion Ratio Lower Upper
154 100
153 118.3 93.1 139.7
152 63.2 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#18
Fluorene
Concen: 0.385 ng
RT: 15.272 min Scan# 1559
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion:166 Resp: 4167
Ion Ratio Lower Upper
166 100
165 100.6 79.5 119.3
167 13.8 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.971 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:188 Resp: 5748
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 5.6 6.2 9.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025




#20
4,6-Dinitro-2-methylphenol
Concen: 0.350 ng
RT: 15.379 min Scan# 1569
Delta R.T. 0.011 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Tgt Ion:198 Resp: 483
Ion Ratio Lower Upper
198 100
51 52.1 51.4 77.0
105 45.9 45.5 68.3





#21
4-Bromophenyl-phenylether
Concen: 0.396 ng
RT: 16.177 min Scan# 1635
Delta R.T. 0.012 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:248 Resp: 162
Ion Ratio Lower Upper
248 100
250 103.0 80.4 120.6
141 31.9 33.3 49.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#22
Hexachlorobenzene
Concen: 0.408 ng
RT: 16.276 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion:284 Resp: 1819
Ion Ratio Lower Upper
284 100
142 34.9 27.0 40.6
249 35.5 28.8 43.2





#23
 Atrazine
 Concen: 0.364 ng
 RT: 16.450 min Scan# 1657
 Delta R.T. 0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC




Tgt Ion:200 Resp: 118
 Ion Ratio Lower Upper
 200 100
 173 25.3 29.2 43.8
 215 53.8 48.8 73.2

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 06/23/2025
 Supervised By :Jagrut Upadhyay 06/24/2025




#24
 Pentachlorophenol
 Concen: 0.363 ng
 RT: 16.624 min Scan# 1671
 Delta R.T. 0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17




Tgt Ion:266 Resp: 805
 Ion Ratio Lower Upper
 266 100
 264 65.6 50.3 75.5
 268 66.2 55.3 82.9





#25
Phenanthrene
Concen: 0.377 ng
RT: 17.009 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:178 Resp: 628
Ion Ratio Lower Upper
178 100
176 19.5 15.2 22.8
179 15.5 12.9 19.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025




#26
Anthracene
Concen: 0.360 ng
RT: 17.108 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17



Tgt Ion:178 Resp: 5522
Ion Ratio Lower Upper
178 100
176 19.0 14.7 22.1
179 16.2 13.0 19.6









#27
Fluoranthene-d10
Concen: 0.414 ng
RT: 19.008 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:212 Resp: 6830
Ion Ratio Lower Upper
212 100
106 5.5 3.0 4.4
104 3.4 2.0 3.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#28
Fluoranthene
Concen: 0.385 ng
RT: 19.036 min Scan# 1936
Delta R.T. -0.005 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17

Tgt Ion:202 Resp: 8102
Ion Ratio Lower Upper
202 100
101 4.5 3.0 4.6
203 16.4 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.162 min Scan# 2116

Delta R.T. -0.009 min

Lab File: BN037365.D


Acq: 21 Jun 2025 01:17

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:240 Resp: 5140

Ion Ratio Lower Upper

240 100

120 9.3 7.5 11.3




236 30.2 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/23/2025

Supervised By :Jagrut Upadhyay 06/24/2025



#30

Pyrene

Concen: 0.388 ng

RT: 19.403 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BN037365.D

Acq: 21 Jun 2025 01:17




Tgt Ion:202 Resp: 8105


Ion Ratio Lower Upper

202 100

200 21.0 16.8 25.2


203 17.7 14.5 21.7





#31
 Terphenyl-d14
 Concen: 0.396 ng
 RT: 19.616 min Scan# 2061
 Delta R.T. 0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC




Tgt Ion:244 Resp: 4639
 Ion Ratio Lower Upper
 244 100
 212 11.1 11.1 16.7
 122 8.6 7.2 10.8

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 06/23/2025
 Supervised By :Jagrut Upadhyay 06/24/2025




#32
 Benzo(a)anthracene
 Concen: 0.376 ng
 RT: 21.153 min Scan# 2273
 Delta R.T. 0.000 min
 Lab File: BN037365.D
 Acq: 21 Jun 2025 01:17



Tgt Ion:228 Resp: 6359
 Ion Ratio Lower Upper
 228 100
 226 27.5 23.0 34.4
 229 20.7 17.4 26.0





#33

Chrysene

Concen: 0.391 ng

RT: 21.198 min Scan# 2119

Delta R.T. -0.009 min

Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:	228	Resp:	801
Ion Ratio	Lower	Upper	
228	100		
226	31.9	25.4	38.2
229	19.9	17.3	25.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/23/2025

Supervised By :Jagrut Upadhyay 06/24/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.385 ng




RT: 21.090 min Scan# 2266


Delta R.T. -0.009 min

Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Tgt Ion:	149	Resp:	2672
Ion Ratio	Lower	Upper	
149	100		
167	31.7	24.6	37.0
279	6.5	6.5	9.7#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.351 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:264 Resp: 568





Ion	Ratio	Lower	Upper
264	100		
260	27.9	21.4	32.2
265	72.3	71.4	107.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/23/2025

Supervised By :Jagrut Upadhyay 06/24/2025



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.382 ng

RT: 25.535 min Scan# 3483




Delta R.T. -0.006 min


Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Tgt Ion:276 Resp: 9725

Ion	Ratio	Lower	Upper
276	100		
138	11.2	2.2	3.2
277	24.1	17.1	25.7





#37

Benzo(b)fluoranthene

Concen: 0.363 ng m

RT: 22.696 min Scan# 2512

Delta R.T. -0.006 min

Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Instrument :

BN037365.D

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:252 Resp: 7592





Ion	Ratio	Lower	Upper
252	100		
253	29.8	26.6	40.0
125	10.0	6.1	9.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/23/2025

Supervised By :Jagrut Upadhyay 06/24/2025



#38

Benzo(k)fluoranthene

Concen: 0.398 ng

RT: 22.740 min Scan# 2527




Delta R.T. -0.006 min


Lab File: BN037365.D

Acq: 21 Jun 2025 01:17

Tgt Ion:252 Resp: 9018


Ion	Ratio	Lower	Upper
252	100		
253	29.0	26.7	40.1
125	8.1	6.5	9.7





#39
Benzo(a)pyrene
Concen: 0.380 ng
RT: 23.255 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:252 Resp: 712
Ion Ratio Lower Upper
252 100
253 34.6 31.6 47.4
125 17.3 8.4 12.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025



#40
Dibenzo(a,h)anthracene
Concen: 0.388 ng
RT: 25.553 min Scan# 3489
Delta R.T. -0.003 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17


Tgt Ion:278 Resp: 7424
Ion Ratio Lower Upper
278 100
139 11.0 10.2 15.4
279 33.0 35.6 53.4





#41
Benzo(g,h,i)perylene
Concen: 0.389 ng
RT: 26.199 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN037365.D
Acq: 21 Jun 2025 01:17

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 885
Ion Ratio Lower Upper
276 100
277 25.1 22.7 34.1
138 13.1 9.4 14.2

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/23/2025
Supervised By :Jagrut Upadhyay 06/24/2025

