

Matrix Spike Summary

Client:	LEO Ingwer, Inc.	SDG No.:	Q2376
Project:	Waste Water 2025	Sample ID:	Q2376-01
Client ID:	EFFLUENTMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.041		0.0016	J	0.04	1	99		06/25/2025

Matrix Spike Summary

Client:	LEO Ingwer, Inc.	SDG No.:	Q2376
Project:	Waste Water 2025	Sample ID:	Q2376-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.042		0.0016	J	0.04	1	101		06/25/2025