

Laboratory Control Sample Summary

Client: LEO Ingwer, Inc.

SDG No.: Q2376

Project: Waste Water 2025

Run No.: LB136260

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168595BS							
Cyanide	mg/L	0.1	0.098		98	1	85-115	06/25/2025