



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: LEO Ingwer, Inc. **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.73	6.0	96	65 - 135	P	06/26/2025	17:28	LB136305
	Chromium	11.3	10.0	113	65 - 135	P	06/26/2025	17:28	LB136305
	Copper	22.8	20.0	114	65 - 135	P	06/26/2025	17:28	LB136305
	Lead	11.9	12.0	99	65 - 135	P	06/26/2025	17:28	LB136305
	Molybdenum	201	200	100	65 - 135	P	06/26/2025	17:28	LB136305
	Nickel	39.3	40.0	98	65 - 135	P	06/26/2025	17:28	LB136305
	Silver	8.68	10.0	87	65 - 135	P	06/26/2025	17:28	LB136305
	Zinc	40.9	40.0	102	65 - 135	P	06/26/2025	17:28	LB136305
CRA	Mercury	0.047	0.05	93	60 - 140	CV	06/27/2025	13:06	LB136313
CRI01	Cadmium	4.81	6.0	80	65 - 135	P	07/03/2025	15:37	LB136377
	Chromium	7.71	10.0	77	65 - 135	P	07/03/2025	15:37	LB136377
	Copper	25.5	20.0	128	65 - 135	P	07/03/2025	15:37	LB136377
	Lead	11.1	12.0	92	65 - 135	P	07/03/2025	15:37	LB136377
	Molybdenum	194	200	97	65 - 135	P	07/03/2025	15:37	LB136377
	Nickel	35.2	40.0	88	65 - 135	P	07/03/2025	15:37	LB136377
	Silver	10.2	10.0	102	65 - 135	P	07/03/2025	15:37	LB136377
	Zinc	38.4	40.0	96	65 - 135	P	07/03/2025	15:37	LB136377