



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	LEO Ingwer, Inc.				SDG No.:	Q2376			
Contract:	LEOI01	Lab Code:	CHEM		Case No.:	Q2376	SAS No.:	Q2376	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB09	Mercury	0.027	+/-0.2	U	0.20	CV	06/27/2025	12:57	LB136313

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.

SDG No.: Q2376

Contract: LEOI01

Lab Code: CHEM

Case No.: Q2376

SAS No.: Q2376

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB27	Mercury	0.027	+/-0.2	U	0.20	CV	06/27/2025	13:01	LB136313
CCB28	Mercury	0.027	+/-0.2	U	0.20	CV	06/27/2025	13:52	LB136313
CCB29	Mercury	0.027	+/-0.2	U	0.20	CV	06/27/2025	14:03	LB136313

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.		SDG No.: Q2376							
Contract:	LEOI01	Lab Code:	CHEM	Case No.:	Q2376	SAS No.:	Q2376		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:23	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:23	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:23	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	17:23	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:23	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:23	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:23	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.		SDG No.: Q2376							
Contract: LEOI01	Lab Code: CHEM	Case No.: Q2376	SAS No.: Q2376						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	17:59	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	17:59	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	17:59	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	17:59	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	17:59	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	17:59	LB136305
CCB02	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	18:52	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	18:52	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	18:52	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	18:52	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	18:52	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	18:52	LB136305
CCB03	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	19:31	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	19:31	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	19:31	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	19:31	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	19:31	LB136305
	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	19:31	LB136305
CCB04	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:07	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:07	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:07	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	20:07	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:07	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:07	LB136305
	Zinc	16.4	+/-20	J	40.0	P	06/26/2025	20:07	LB136305
CCB05	Cadmium	0.50	+/-3	U	6.00	P	06/26/2025	20:51	LB136305
	Chromium	2.12	+/-5	U	10.0	P	06/26/2025	20:51	LB136305
	Copper	4.60	+/-10	U	20.0	P	06/26/2025	20:51	LB136305
	Lead	2.30	+/-6	U	12.0	P	06/26/2025	20:51	LB136305
	Molybdenum	18.7	+/-100	U	200	P	06/26/2025	20:51	LB136305
	Nickel	3.06	+/-20	U	40.0	P	06/26/2025	20:51	LB136305
	Silver	1.62	+/-5	U	10.0	P	06/26/2025	20:51	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.

SDG No.: Q2376

Contract: LEOI01

Lab Code: CHEM

Case No.: Q2376

SAS No.: Q2376

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	3.50	+/-20	U	40.0	P	06/26/2025	20:51	LB136305

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.		SDG No.: Q2376							
Contract:	LEOI01	Lab Code:	CHEM	Case No.:	Q2376	SAS No.:	Q2376		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	15:32	LB136377
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	15:32	LB136377
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	15:32	LB136377
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	15:32	LB136377
	Molybdenum	18.7	+/-100	U	200	P	07/03/2025	15:32	LB136377
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	15:32	LB136377
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	15:32	LB136377
	Zinc	3.50	+/-20	U	40.0	P	07/03/2025	15:32	LB136377

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LEO Ingwer, Inc.		SDG No.: Q2376							
Contract: LEOI01	Lab Code: CHEM	Case No.: Q2376	SAS No.: Q2376						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	16:10	LB136377
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	16:10	LB136377
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	16:10	LB136377
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	16:10	LB136377
	Molybdenum	18.7	+/-100	U	200	P	07/03/2025	16:10	LB136377
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	16:10	LB136377
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	16:10	LB136377
	Zinc	3.50	+/-20	U	40.0	P	07/03/2025	16:10	LB136377
CCB02	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	16:57	LB136377
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	16:57	LB136377
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	16:57	LB136377
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	16:57	LB136377
	Molybdenum	18.7	+/-100	U	200	P	07/03/2025	16:57	LB136377
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	16:57	LB136377
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	16:57	LB136377
	Zinc	3.50	+/-20	U	40.0	P	07/03/2025	16:57	LB136377
CCB03	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	17:37	LB136377
	Chromium	2.12	+/-5	U	10.0	P	07/03/2025	17:37	LB136377
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	17:37	LB136377
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	17:37	LB136377
	Molybdenum	18.7	+/-100	U	200	P	07/03/2025	17:37	LB136377
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	17:37	LB136377
	Silver	1.62	+/-5	U	10.0	P	07/03/2025	17:37	LB136377
	Zinc	3.50	+/-20	U	40.0	P	07/03/2025	17:37	LB136377