

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: LEO Ingwer, Inc. **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-1.09	1.0	109	-5	7	06/26/2025	17:37	LB136305
	Chromium	50.8	52.0	98	42	62	06/26/2025	17:37	LB136305
	Copper	-5.23	2.0	262	-18	22	06/26/2025	17:37	LB136305
	Lead	2.20			-12	12	06/26/2025	17:37	LB136305
	Molybdenum	-10.4			0	0	06/26/2025	17:37	LB136305
	Nickel	3.31	2.0	166	-38	42	06/26/2025	17:37	LB136305
	Silver	2.97			-10	10	06/26/2025	17:37	LB136305
	Zinc	12.0			-40	40	06/26/2025	17:37	LB136305
ICSAB01	Cadmium	970	970	100	826	1120	06/26/2025	17:42	LB136305
	Chromium	541	540	100	460	624	06/26/2025	17:42	LB136305
	Copper	468	510	92	434	588	06/26/2025	17:42	LB136305
	Lead	49.4	49.0	101	37	61	06/26/2025	17:42	LB136305
	Molybdenum	933			0	0	06/26/2025	17:42	LB136305
	Nickel	947	950	100	810	1100	06/26/2025	17:42	LB136305
	Silver	211	200	106	170	232	06/26/2025	17:42	LB136305
	Zinc	1010	950	106	809	1095	06/26/2025	17:42	LB136305
ICSA	Cadmium	2.45	1.0	245	-5	7	06/26/2025	17:46	LB136305
	Chromium	37.4	52.0	72	42	62	06/26/2025	17:46	LB136305
	Copper	-13.9	2.0	695	-18	22	06/26/2025	17:46	LB136305
	Lead	35.9			-12	12	06/26/2025	17:46	LB136305
	Molybdenum	-56.7			0	0	06/26/2025	17:46	LB136305
	Nickel	-1.58	2.0	79	-38	42	06/26/2025	17:46	LB136305
	Silver	18.8			-10	10	06/26/2025	17:46	LB136305
	Zinc	16.0			-40	40	06/26/2025	17:46	LB136305
ICSAB	Cadmium	871	970	90	826	1120	06/26/2025	17:50	LB136305
	Chromium	530	540	98	460	624	06/26/2025	17:50	LB136305
	Copper	479	510	94	434	588	06/26/2025	17:50	LB136305
	Lead	70.9	49.0	145	37	61	06/26/2025	17:50	LB136305
	Molybdenum	840			0	0	06/26/2025	17:50	LB136305
	Nickel	864	950	91	810	1100	06/26/2025	17:50	LB136305
	Silver	185	200	92	170	232	06/26/2025	17:50	LB136305
	Zinc	959	950	101	809	1095	06/26/2025	17:50	LB136305
ICSA01	Cadmium	-0.62	1.0	62	-5	7	07/03/2025	15:41	LB136377
	Chromium	53.3	52.0	102	42	62	07/03/2025	15:41	LB136377
	Copper	14.6	2.0	730	-18	22	07/03/2025	15:41	LB136377
	Lead	1.07			-12	12	07/03/2025	15:41	LB136377
	Molybdenum	-6.48			0	0	07/03/2025	15:41	LB136377
	Nickel	13.6	2.0	680	-38	42	07/03/2025	15:41	LB136377
	Silver	4.00			-10	10	07/03/2025	15:41	LB136377
	Zinc	12.6			-40	40	07/03/2025	15:41	LB136377
ICSAB01	Cadmium	946	970	98	826	1120	07/03/2025	15:45	LB136377

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: LEO Ingwer, Inc. **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Chromium	536	540	99	460	624	07/03/2025	15:45	LB136377
	Copper	465	510	91	434	588	07/03/2025	15:45	LB136377
	Lead	44.5	49.0	91	37	61	07/03/2025	15:45	LB136377
	Molybdenum	1030			0	0	07/03/2025	15:45	LB136377
	Nickel	940	950	99	810	1100	07/03/2025	15:45	LB136377
	Silver	230	200	115	170	232	07/03/2025	15:45	LB136377
	Zinc	1010	950	106	809	1095	07/03/2025	15:45	LB136377
ICSA	Cadmium	-7.06	1.0	706	-5	7	07/03/2025	15:50	LB136377
	Chromium	56.8	52.0	109	42	62	07/03/2025	15:50	LB136377
	Copper	130	2.0	6500	-18	22	07/03/2025	15:50	LB136377
	Lead	53.6			-12	12	07/03/2025	15:50	LB136377
	Molybdenum	-40.7			0	0	07/03/2025	15:50	LB136377
	Nickel	11.6	2.0	580	-38	42	07/03/2025	15:50	LB136377
	Silver	-0.40			-10	10	07/03/2025	15:50	LB136377
	Zinc	228			-40	40	07/03/2025	15:50	LB136377
ICSAB	Cadmium	882	970	91	826	1120	07/03/2025	15:54	LB136377
	Chromium	539	540	100	460	624	07/03/2025	15:54	LB136377
	Copper	574	510	112	434	588	07/03/2025	15:54	LB136377
	Lead	87.4	49.0	178	37	61	07/03/2025	15:54	LB136377
	Molybdenum	991			0	0	07/03/2025	15:54	LB136377
	Nickel	897	950	94	810	1100	07/03/2025	15:54	LB136377
	Silver	223	200	112	170	232	07/03/2025	15:54	LB136377
	Zinc	974	950	102	809	1095	07/03/2025	15:54	LB136377