

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LEO Ingwer, Inc. **Level:** LOW **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
Matrix: Water **Sample ID:** Q2376-02 **Client ID:** EFFLUENT-METALSDUP
Percent Solids for Sample: NA **Duplicate ID** Q2376-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.17	J	0.11	J	44		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LEO Ingwer, Inc. **Level:** LOW **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
Matrix: Water **Sample ID:** Q2376-02MS **Client ID:** EFFLUENT-METALSMSD
Percent Solids for Sample: NA **Duplicate ID** Q2376-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.89		3.77		3		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LEO Ingwer, Inc. **Level:** LOW **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
Matrix: Water **Sample ID:** Q2404-12 **Client ID:** 1ST-FL-CHILD-CARE-SINK-2DUP
Percent Solids for Sample: NA **Duplicate ID** Q2404-12DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	1.02	J	1.33	J	26		P
Copper	ug/L	20	71.5		71.9		1		P
Lead	ug/L	20	2.58	J	2.95	J	13		P
Molybdenum	ug/L	20	100	U	100	U			P
Nickel	ug/L	20	20.0	U	20.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Zinc	ug/L	20	55.9		56.1		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LEO Ingwer, Inc. **Level:** LOW **SDG No.:** Q2376
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** Q2376 **SAS No.:** Q2376
Matrix: Water **Sample ID:** Q2404-12MS **Client ID:** 1ST-FL-CHILD-CARE-SINK-2MSD
Percent Solids for Sample: NA **Duplicate ID** Q2404-12MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	91.8		91.0		1		P
Chromium	ug/L	20	205		203		1		P
Copper	ug/L	20	233		234		0		P
Lead	ug/L	20	464		460		1		P
Molybdenum	ug/L	20	209		208		0		P
Nickel	ug/L	20	236		234		1		P
Silver	ug/L	20	30.0		30.2		1		P
Zinc	ug/L	20	159		159		0		P