

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LEO Ingwer, Inc.

SDG No.: Q2376

Contract: LEOI01

Lab Code: CHEM

Case No.: Q2376

SAS No.: Q2376

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168611BS							
Cadmium	ug/L	100	91.5		92	85 - 115	P
Chromium	ug/L	200	198		99	85 - 115	P
Copper	ug/L	150	154		103	85 - 115	P
Lead	ug/L	500	456		91	85 - 115	P
Molybdenum	ug/L	200	191		96	85 - 115	P
Nickel	ug/L	250	233		93	85 - 115	P
Silver	ug/L	37.5	34.5		92	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LEO Ingwer, Inc.

SDG No.: Q2376

Contract: LEOI01

Lab Code: CHEM

Case No.: Q2376

SAS No.: Q2376

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168650BS Mercury	ug/L	4.0	3.89		97	85 - 115	CV