

## **ANALYTICAL RESULTS SUMMARY**

SEMI-VOLATILE ORGANICS  
VOLATILE ORGANICS

**PROJECT NAME : FORMER SCHLUMBERGER STC PTC SITE D3868221**

**JACOBS ENGINEERING GROUP, INC.**

**412 Mt. Kemble Ave**

**Downtown Building**

**Morristown, NJ - 07960**

**Phone No: 9732670555**

**ORDER ID : Q2377**

**ATTENTION : John Ynfante**



**Laboratory Certification ID # 20012**



1) Signature Page	3
2) Case Narrative	5
2.1) VOCMS Group3- Case Narrative	5
2.2) VOC-SIM- Case Narrative	7
2.3) SVOC-SIMGroup1- Case Narrative	9
3) Qualifier Page	11
4) QA Checklist	12
5) VOCMS Group3 Data	13
6) VOC-SIM Data	18
7) SVOC-SIMGroup1 Data	22
8) Shipping Document	26
8.1) CHAIN OF CUSTODY	27
8.2) Lab Certificate	28
8.3) Internal COC	29

1
2
3
4
5
6
7
8

# DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

1

Laboratory Name : Alliance Technical Group LLC

Client : JACOBS Engineering Group, Inc.

Project Location : Princeton Junction

Project Number : D3868221

Laboratory Sample ID(s) : Q2377

Sampling Date(s) : 6/19/2025

List DKQP Methods Used (e.g., 8260,8270, et Cetra) **524.2,8270-Modified,SFAM\_VOCSIM,SOP**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1A	Were the method specified handling, preservation, and holding time requirements met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
3	Were samples received at an appropriate temperature (4±2° C)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt?  b)Were these reporting limits met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No  <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

## Cover Page

**Order ID :** Q2377

**Project ID :** Former Schlumberger STC PTC Site D3868221

**Client :** JACOBS Engineering Group, Inc.

**Lab Sample Number**

Q2377-01  
Q2377-02  
Q2377-03

**Client Sample Number**

PW-B6-L66-061925  
PW-B6-L66-061925-SIM  
TB01-061925

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 2:36 pm, Jul 07, 2025*

Date: 7/7/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site D3868221**

**Project # N/A**

**Order ID # Q2377**

**Test Name: VOCMS Group3**

### **A. Number of Samples and Date of Receipt:**

3 Water samples were received on 06/19/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1, VOC-SIM and VOCMS Group3. This data package contains results for VOCMS Group3.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_U were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group3 was based on method 524.2.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for { VU0624WBSD01 } with File ID: VU063433.D met criteria except for 1,1-Dichloroethene[26%] due to difference in results of BS and BSD.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

### **E. Additional Comments:**

This data package has been revised due to parameter list changed as per client request.

Samples for MS/MSD for VOC analysis were not provided with this set of samples.

The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 2:36 pm, Jul 07, 2025*



## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site D3868221**

**Project # N/A**

**Order ID # Q2377**

**Test Name: VOC-SIM**

### **A. Number of Samples and Date of Receipt:**

3 Water samples were received on 06/19/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOC-SIM Group1, VOC-SIM and VOCMS Group3. This data package contains results for VOC-SIM.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_V were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOC-SIM was based on method SFAM\_VOCSIM.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

### **E. Additional Comments:**

### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 2:36 pm, Jul 07, 2025*

Signature\_\_\_\_\_

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site D3868221**

**Project # N/A**

**Order ID # Q2377**

**Test Name: SVOC-SIMGroup1**

### **A. Number of Samples and Date of Receipt:**

3 Water samples were received on 06/19/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1, VOC-SIM and VOCMS Group3. This data package contains results for SVOC-SIMGroup1.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Tuning criteria met requirements.

### **E. Additional Comments:**

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 2:37 pm, Jul 07, 2025*

**DATA REPORTING QUALIFIERS- ORGANIC**

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q2377

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MOHAMMAD AHMED

Date: 07/07/2025

**Hit Summary Sheet**  
SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



# SAMPLE DATA

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/19/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/19/25
Client Sample ID:	PW-B6-L66-061925	SDG No.:	Q2377
Lab Sample ID:	Q2377-01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VU063434.D	1	06/24/25 12:54	VU062425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.13	U	0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	0.12	U	0.12	0.50	ug/L
75-34-3	1,1-Dichloroethane	0.13	U	0.13	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	0.13	U	0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	0.12	U	0.12	0.50	ug/L
71-43-2	Benzene	0.11	U	0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	0.16	U	0.16	0.50	ug/L
79-01-6	Trichloroethene	0.13	U	0.13	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	0.13	U	0.13	0.50	ug/L
127-18-4	Tetrachloroethene	0.14	U	0.14	0.50	ug/L
<b>SURROGATES</b>						
2199-69-1	1,2-Dichlorobenzene-d4	0.80		70 (70) - 130 (130)	80%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.80		70 (70) - 130 (130)	80%	SPK: 1
<b>INTERNAL STANDARDS</b>						
462-06-6	Fluorobenzene	46300	6.107			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/19/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/19/25
Client Sample ID:	TB01-061925	SDG No.:	Q2377
Lab Sample ID:	Q2377-03	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VU063435.D	1	06/24/25 13:22	VU062425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.13	U	0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	0.12	U	0.12	0.50	ug/L
75-34-3	1,1-Dichloroethane	0.13	U	0.13	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	0.13	U	0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	0.12	U	0.12	0.50	ug/L
71-43-2	Benzene	0.11	U	0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	0.16	U	0.16	0.50	ug/L
79-01-6	Trichloroethene	0.13	U	0.13	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	0.13	U	0.13	0.50	ug/L
127-18-4	Tetrachloroethene	0.14	U	0.14	0.50	ug/L
<b>SURROGATES</b>						
2199-69-1	1,2-Dichlorobenzene-d4	0.92		70 (70) - 130 (130)	92%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.89		70 (70) - 130 (130)	89%	SPK: 1
<b>INTERNAL STANDARDS</b>						
462-06-6	Fluorobenzene	40700	6.107			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## LAB CHRONICLE

<b>OrderID:</b>	Q2377	<b>OrderDate:</b>	6/20/2025 11:23:00 AM
<b>Client:</b>	JACOBS Engineering Group, Inc.	<b>Project:</b>	Former Schlumberger STC PTC Site D3868221
<b>Contact:</b>	John Ynfante	<b>Location:</b>	D51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2377-01</b>	<b>PW-B6-L66-061925</b>	<b>Water</b>	VOCMS Group3	524.2	<b>06/19/25</b>		06/24/25	<b>06/19/25</b>
<b>Q2377-02</b>	<b>PW-B6-L66-061925-S IM</b>	<b>Water</b>	VOC-SIM	SFAM_VOCSI M	<b>06/19/25</b>		06/23/25	<b>06/19/25</b>
<b>Q2377-03</b>	<b>TB01-061925</b>	<b>Water</b>	VOCMS Group3	524.2	<b>06/19/25</b>		06/24/25	<b>06/19/25</b>

**Hit Summary Sheet**  
SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



# SAMPLE DATA

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/19/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/19/25
Client Sample ID:	PW-B6-L66-061925-SIM	SDG No.:	Q2377
Lab Sample ID:	Q2377-02	Matrix:	Water
Analytical Method:	SFAM_VOCSIM	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOC-SIM
GC Column:	DB-624UI ID : 0.18	Level :	
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VV038820.D	1		06/23/25 10:00	VV062325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl chloride	0.021	U	0.021	0.050	ug/L
<b>SURROGATES</b>						
6745-35-3	Vinyl Chloride-d3	0.42		40 - 130	84.6%	SPK: 0.5
17060-07-0	1,2-Dichloroethane-d4	0.43		70 - 130	85.2%	SPK: 0.5
93952-08-0	1,2-Dichloropropane-d6	0.47		60 - 140	93.8%	SPK: 0.5
2037-26-5	Toluene-d8	0.46		70 - 130	92%	SPK: 0.5
33685-54-0	1,1,2,2-Tetrachloroethane-d2	0.42		65 - 120	84.4%	SPK: 0.5
<b>INTERNAL STANDARDS</b>						
3114-55-4	Chlorobenzene-d5	6820	8.785			
540-36-3	1,4-Difluorobenzene	7440	5.566			
3855-82-1	1,4-Dichlorobenzene-d4	2940	11.191			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## LAB CHRONICLE

<b>OrderID:</b>	Q2377	<b>OrderDate:</b>	6/20/2025 11:23:00 AM
<b>Client:</b>	JACOBS Engineering Group, Inc.	<b>Project:</b>	Former Schlumberger STC PTC Site D3868221
<b>Contact:</b>	John Ynfante	<b>Location:</b>	D51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2377-01</b>	<b>PW-B6-L66-061925</b>	<b>Water</b>	VOCMS Group3	524.2	<b>06/19/25</b>		06/24/25	<b>06/19/25</b>
<b>Q2377-02</b>	<b>PW-B6-L66-061925-S IM</b>	<b>Water</b>	VOC-SIM	SFAM_VOCSI M	<b>06/19/25</b>		06/23/25	<b>06/19/25</b>
<b>Q2377-03</b>	<b>TB01-061925</b>	<b>Water</b>	VOCMS Group3	524.2	<b>06/19/25</b>		06/24/25	<b>06/19/25</b>



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q2377  
**Client:** JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :		0.00			
			Total Concentration:		0.00			



# SAMPLE DATA

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/19/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/19/25
Client Sample ID:	PW-B6-L66-061925	SDG No.:	Q2377
Lab Sample ID:	Q2377-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037362.D	1	06/20/25 12:03	06/20/25 22:52	PB168563

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
123-91-1	1,4-Dioxane	0.070	U	0.070	0.20	ug/L
<b>SURROGATES</b>						
7297-45-2	2-Methylnaphthalene-d10	0.33		30 (20) - 150 (139)	83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 (54) - 150 (157)	94%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		30 (27) - 130 (154)	77%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		30 (30) - 130 (155)	88%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		30 (54) - 130 (175)	105%	SPK: 0.4
<b>INTERNAL STANDARDS</b>						
3855-82-1	1,4-Dichlorobenzene-d4	1840	7.568			
1146-65-2	Naphthalene-d8	4230	10.351			
15067-26-2	Acenaphthene-d10	2930	14.213			
1517-22-2	Phenanthrene-d10	6080	16.971			
1719-03-5	Chrysene-d12	5240	21.162			
1520-96-3	Perylene-d12	5350	23.348			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q2377	OrderDate:	6/20/2025 11:23:00 AM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	Location:	D51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2377-01	PW-B6-L66-061925	Water	SVOC-SIMGroup1	8270-Modified	06/19/25	06/20/25	06/20/25	06/19/25



# SHIPPING DOCUMENTS

# CHEMTECH

## CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092  
(908) 789-8900 • Fax (908) 789-8922  
www.chemtech.net

CHEMTECH PROJECT NO. **Q2377**  
QUOTE NO. **Q2377**  
COC Number **2039472**

8  
8.1

### CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Jacobs**  
ADDRESS: **412 Mt. Kemble Ave., Ste. 100**  
CITY: **Morrisstown** STATE: **NJ** ZIP: **07960**  
ATTENTION: **John Infante John.Infante@Jacobs.com**  
PHONE: FAX:

### CLIENT PROJECT INFORMATION

PROJECT NAME: **STC Princeton**  
PROJECT NO.: **D3868221** LOCATION: **Princeton Junction**  
PROJECT MANAGER: **Mary Murphy**  
e-mail: **Mary.Murphy@Jacobs.com**  
PHONE: FAX:

### CLIENT BILLING INFORMATION

BILL TO: **Mary Murphy** PO#:  
ADDRESS:  
CITY STATE: ZIP:  
ATTENTION: PHONE:

### ANALYSIS

### DATA TURNAROUND INFORMATION

FAX (RUSH) **\*RUSH TAT (2 Day) \*** DAYS\*  
HARDCOPY (DATA PACKAGE): DAYS\*  
EDD: DAYS\*

\*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

### DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP  
☒ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other  
☐ EDD FORMAT

1 Site specific VOCs  
2 1,4 Dioxane (8270-SIM)  
3 VOCs (SPAM01.1-SIM)  
4 LVC only

### PRESERVATIVES

### COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		A/E	E	A/E							
1.	PW-B6-L66-061925	AQ		X	6/19/25	11:00	5	X	X								
2.	PW-B6-L66-061925-SIM	AQ		X		11:00	3			X							
3.	PW-B6-L66-061925-FD <sup>AM 6/19/25</sup>	AQ		X		11:05	5	X	X								
4.	PW-B6-L66-061925-SIM-FD <sup>AM 6/19/25</sup>	AQ		X		11:05	3			X							
5.	TBO1-061925	D1		X			2	X									
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP
1. <i>[Signature]</i>	6/19/25 / 17:10	1. <i>[Signature]</i>	Comments: See work order for list of site specific VOCs
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3.		3.	

Page 1 of 1

CLIENT: ☒ Hand Delivered ☐ Other  
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete  
☐ YES ☐ NO

### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

## LOGIN REPORT/SAMPLE TRANSFER

<b>Order ID :</b> Q2377	JACO05	<b>Order Date :</b> 6/20/2025 11:23:00 AM	<b>Project Mgr :</b>
<b>Client Name :</b> JACOBS Engineering Grou		<b>Project Name :</b> Former Schlumberger STC	<b>Report Type :</b> Level 3
<b>Client Contact :</b> John Ynfante		<b>Receive DateTime :</b> 6/19/2025 5:10:00 PM	<b>EDD Type :</b> CH2MHILL
<b>Invoice Name :</b> JACOBS Engineering Grou		<b>Purchase Order :</b>	<b>Hard Copy Date :</b>
<b>Invoice Contact :</b> John Ynfante			<b>Date Signoff :</b>

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2377-01	PW-B6-L66-061925	Water	06/19/2025	11:00	VOCMS Group3		524.2 <del>8260-Low</del>		2 Bus. Days
Q2377-02	PW-B6-L66-061925-SIM	Water	06/19/2025	11:00	VOC-SIM		SFAM_VOCSIM		2 Bus. Days
Q2377-03	TB01-061925	Water	06/19/2025	11:00	VOCMS Group3		524.2 <del>8260-Low</del>		2 Bus. Days

Relinquished By : af

Date / Time : 6/20/25 1150

Received By : Samm

Date / Time : 06/20/25

Storage Area : VOA Refridgerator Room