

Surrogate Summary

SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168563BL	PB168563BL	2-Methylnaphthalene-d10	0.4	0.32	81		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	93		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.30	75		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.33	81		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.37	93		30 (54)	130 (175)
PB168563BS	PB168563BS	2-Methylnaphthalene-d10	0.4	0.49	123		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	86		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.37	92		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.39	97		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.39	98		30 (54)	130 (175)
PB168563BSD	PB168563BSD	2-Methylnaphthalene-d10	0.4	0.39	97		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	86		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.39	97		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.40	99		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.38	94		30 (54)	130 (175)
Q2377-01	PW-B6-L66-061925	2-Methylnaphthalene-d10	0.4	0.33	83		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.38	94		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.31	77		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	88		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.42	105		30 (54)	130 (175)