

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified **DataFile:** BN037363.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168563BS	1,4-Dioxane	0.4	0.32	ug/L	80				20 (65)	160 (116)	