

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified **DataFile:** BN037364.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High RPD
PB168563BSD	1,4-Dioxane	0.4	0.30	ug/L	75	6			20 (65)	160 (116) 20 (27)