

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2377 SAS No.: Q2377 SDG No.: Q2377

Instrument ID: MSVOA_V Calibration Date/Time: 06/23/2025 08:44

Lab File ID: VV038818.D Init. Calib. Date(s): 06/16/2025 06/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:28 11:36

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRFCAL	MIN RRF	%D	MAX%D
Vinyl chloride	0.900	0.916	0.01	1.7	30
Vinyl Chloride-d3	0.475	0.483	0.01	1.7	30
1,2-Dichloroethane-d4	0.233	0.238	0.01	2.2	25
1,2-Dichloropropane-d6	0.283	0.311	0.1	9.9	20
Toluene-d8	0.572	0.624	0.2	9.2	20
1,1,2,2-Tetrachloroethane-d2	0.199	0.217	0.01	9.1	25

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.