

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2377 SAS No.: Q2377 SDG No.: Q2377

Instrument ID: MSVOA_V Calibration Date/Time: 06/23/2025 13:37

Lab File ID: VV038821.D Init. Calib. Date(s): 06/16/2025 06/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:28 11:36

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRFCAL	MIN RRF	%D	MAX%D
Vinyl chloride	0.900	0.907	0.01	0.8	50
Vinyl Chloride-d3	0.475	0.494	0.01	4	50
1,2-Dichloroethane-d4	0.233	0.248	0.01	6.4	50
1,2-Dichloropropane-d6	0.283	0.332	0.1	17.3	50
Toluene-d8	0.572	0.623	0.2	9	50
1,1,2,2-Tetrachloroethane-d2	0.199	0.224	0.01	12.5	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.