

Surrogate Summary

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Analytical Method: SWSFAM_VOCSIM

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2377-02	PW-B6-L66-061925-SIM	Vinyl chloride-d3	0.5	0.42	85	40	130
		1,2-Dichloroethane-d4	0.5	0.43	85	70	130
		1,2-Dichloropropane-d6	0.5	0.47	94	60	140
		Toluene-d8	0.5	0.46	92	70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.42	84	65	120
VV0623WBL01	VBLK229	Vinyl chloride-d3	0.5	0.48	95	40	130
		1,2-Dichloroethane-d4	0.5	0.49	98	70	130
		1,2-Dichloropropane-d6	0.5	0.53	107	60	140
		Toluene-d8	0.5	0.51	102	70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.51	102	65	120