

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JAC005
 Lab Code: CHEM Case No.: Q2377 SAS No.: Q2377 SDG NO.: Q2377
 Lab File ID: VV038818.D Date Analyzed: 06/23/2025
 Instrument ID: MSVOA_V Time Analyzed: 08:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 (DFB) AREA #	RT #	IS2 (CBZ) AREA #	RT #	IS3 (DCB) AREA #	RT #
12 HOUR STD	6318	5.57	5798	8.79	2580	11.19
UPPER LIMIT	12636	5.736	11596	8.955	5160	11.361
LOWER LIMIT	3159	5.396	2899	8.615	1290	11.021
EPA SAMPLE NO.						
PW-B6-L66-061925-SIM	7436	5.57	6823	8.79	2936	11.19
VBLK229	6610	5.57	6099	8.79	2691	11.19

IS1 (DFB) = 1,4-Difluorobenzene

IS2 (CBZ) = Chlorobenzene-d5

IS3 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.