

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062425\
 Data File : VU063430.D
 Acq On : 24 Jun 2025 10:17
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jun 25 06:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U062325DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Jun 25 03:53:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.384	0.347	9.6	96	0.00
3 t	Chloromethane	0.506	0.382	24.5	85	0.00
4 Rt	Vinyl Chloride	0.433	0.334	22.9	82	0.00
5 T	Bromomethane	0.208	0.167	19.7	91	0.00
6 T	Chloroethane	0.257	0.175	31.9#	73	0.00
7 T	Trichlorofluoromethane	0.494	0.399	19.2	86	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.263	0.209	20.5	85	0.00
9 Rt	1,1-Dichloroethene	0.263	0.210	20.2	85	0.00
10 t	Iodomethane	0.212	0.240	-13.2	106	0.00
11 t	Allyl Chloride	0.468	0.301	35.7#	68	0.00
12 t	Acrylonitrile	0.076	0.059	22.4	84	0.00
13 T	Acetone	0.076	0.035	53.9#	54	0.00
14 T	Carbon Disulfide	0.916	0.681	25.7	80	0.00
15 RT	Methylene Chloride	0.333	0.254	23.7	84	0.00
16 RT	trans-1,2-Dichloroethene	0.295	0.258	12.5	94	0.00
17 t	1,1-Dichloroethane	0.620	0.590	4.8	102	0.00
18 T	2-Butanone	0.106	0.088	17.0	90	0.00
19	Cyclohexane	0.469	0.469	0.0	105	0.00
20	Methylcyclohexane	0.436	0.430	1.4	93	0.00
21 T	2,2-Dichloropropane	0.441	0.408	7.5	100	0.00
22 RT	cis-1,2-Dichloroethene	0.322	0.302	6.2	102	0.00
23 t	Diethyl Ether	0.237	0.157	33.8#	70	0.00
24 t	tert-Butyl Alcohol	0.025	0.019	24.0	80	0.00
25 t	Methyl tert-Butyl Ether	0.714	0.623	12.7	91	0.00
26 t	Bromochloromethane	0.126	0.116	7.9	98	0.00
27 t	Chloroform	0.582	0.538	7.6	100	0.00
28 RT	1,1,1-Trichloroethane	0.472	0.449	4.9	98	0.00
29 T	1,1-Dichloropropene	0.431	0.416	3.5	100	0.00
30 RT	Carbon Tetrachloride	0.377	0.366	2.9	99	0.00
31 t	Isopropyl Ether	0.969	0.916	5.5	99	0.00
32	Ethyl-t-butyl ether	0.821	0.774	5.7	99	0.00
33	Tert-Amyl methyl ether	0.588	0.594	-1.0	100	0.00
34 t	Propionitrile	0.031	0.027	12.9	102	0.00
35 RT	Benzene	1.303	1.270	2.5	100	0.00
36 RT	1,2-Dichloroethane	0.381	0.350	8.1	96	0.00
37 RT	Trichloroethene	0.308	0.291	5.5	98	0.00
38 Rt	1,2-Dichloropropane	0.360	0.358	0.6	101	0.00
39 t	Methacrylonitrile	0.127	0.111	12.6	94	0.00
40 t	Methyl acrylate	0.179	0.159	11.2	85	0.00
41 t	Tetrahydrofuran	0.069	0.059	14.5	98	0.00
42 t	1-Chlorobutane	0.611	0.596	2.5	101	0.00
43 T	Dibromomethane	0.168	0.161	4.2	100	0.00
44 T	Bromodichloromethane	0.410	0.388	5.4	100	0.00
45 T	4-Methyl-2-Pentanone	0.180	0.178	1.1	94	0.00
46 t	t-1,4-Dichloro-2-butene	0.059	0.047	20.3	89	0.00
47 t	Methyl methacrylate	0.130	0.142	-9.2	97	0.00
48 t	Ethyl methacrylate	0.233	0.258	-10.7	98	0.00
49 Rt	Toluene	0.715	0.757	-5.9	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062425\
 Data File : VU063430.D
 Acq On : 24 Jun 2025 10:17
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jun 25 06:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U062325DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Jun 25 03:53:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.295	0.327	-10.8	104	0.00
51 T	cis-1,3-Dichloropropene	0.397	0.419	-5.5	101	0.00
52 RT	1,1,2-Trichloroethane	0.246	0.231	6.1	98	0.00
53 t	1,3-Dichloropropane	0.423	0.409	3.3	98	0.00
54 t	2-Hexanone	0.121	0.116	4.1	90	0.00
55 t	Dibromochloromethane	0.254	0.250	1.6	100	0.00
56 T	1,2-Dibromoethane	0.199	0.188	5.5	100	0.00
57 S	4-Bromofluorobenzene	0.367	0.409	-11.4	105	0.00
58 RT	Tetrachloroethene	0.302	0.289	4.3	97	0.00
59 Rt	Chlorobenzene	0.766	0.763	0.4	99	0.00
60 T	1,1,1,2-Tetrachloroethane	0.276	0.266	3.6	100	0.00
61 t	Pentachloroethane	0.225	0.214	4.9	103	0.00
62 t	Hexachloroethane	0.202	0.201	0.5	103	0.00
63 Rt	Ethyl Benzene	1.228	1.352	-10.1	101	0.00
64 RT	m/p-Xylenes	0.479	0.546	-14.0	101	0.00
65 RT	o-Xylene	0.466	0.500	-7.3	99	0.00
66 RT	Styrene	0.759	0.872	-14.9	101	0.00
67 t	Bromoform	0.139	0.132	5.0	97	0.00
68 S	1,2-Dichlorobenzene-d4	0.369	0.351	4.9	96	0.00
69 T	Isopropylbenzene	1.081	1.175	-8.7	100	0.00
70 T	1,1,2,2-Tetrachloroethane	0.314	0.292	7.0	98	0.00
71 T	1,2,3-Trichloropropane	0.257	0.249	3.1	113	0.00
72 t	Bromobenzene	0.303	0.303	0.0	98	0.00
73 t	n-propylbenzene	0.328	0.360	-9.8	100	0.00
74 t	2-Chlorotoluene	0.300	0.320	-6.7	100	0.00
75 t	1,3,5-Trimethylbenzene	1.051	1.204	-14.6	100	0.00
76 t	4-Chlorotoluene	0.314	0.336	-7.0	102	0.00
77 t	tert-Butylbenzene	0.993	1.071	-7.9	100	0.00
78 t	1,2,4-Trimethylbenzene	1.058	1.233	-16.5	102	0.00
79 t	sec-Butylbenzene	1.406	1.541	-9.6	101	0.00
80	Nitrobenzene	0.014	0.012	14.3	90	0.00
81 t	p-Isopropyltoluene	1.099	1.257	-14.4	100	0.00
82 t	1,3-Dichlorobenzene	0.628	0.621	1.1	99	0.00
83 Rt	1,4-Dichlorobenzene	0.652	0.649	0.5	100	0.00
84 t	n-Butylbenzene	1.151	1.235	-7.3	98	0.00
85 Rt	1,2-Dichlorobenzene	0.601	0.591	1.7	98	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.043	0.043	0.0	96	0.00
87 Rt	1,2,4-Trichlorobenzene	0.335	0.323	3.6	94	0.00
88 t	Hexachlorobutadiene	0.238	0.228	4.2	98	0.00
89 t	Naphthalene	0.558	0.525	5.9	85	0.00
90 t	1,2,3-Trichlorobenzene	0.333	0.321	3.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0