

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2377

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW524.2

**Datafile :** VU063432.D

| Lab Sample ID | Parameter                 | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|---------------------------|-------|--------|------|-----|-----|------|---------|----------------|-----|
| VU0624WBS01   | Vinyl Chloride            | 2     | 1.70   | ug/L | 85  |     |      | 70 (70) | 130 (130)      |     |
|               | 1,1-Dichloroethene        | 2     | 1.70   | ug/L | 85  |     |      | 70 (70) | 130 (130)      |     |
|               | 1,1-Dichloroethane        | 2     | 2.00   | ug/L | 100 |     |      | 70 (70) | 130 (130)      |     |
|               | cis-1,2-Dichloroethene    | 2     | 2.00   | ug/L | 100 |     |      | 70 (70) | 130 (130)      |     |
|               | Benzene                   | 2     | 2.00   | ug/L | 100 |     |      | 70 (70) | 130 (130)      |     |
|               | 1,2-Dichloroethane        | 2     | 2.00   | ug/L | 100 |     |      | 70 (70) | 130 (130)      |     |
|               | Trichloroethene           | 2     | 1.90   | ug/L | 95  |     |      | 70 (70) | 130 (130)      |     |
|               | 1,1,2-Trichloroethane     | 2     | 1.90   | ug/L | 95  |     |      | 70 (70) | 130 (130)      |     |
|               | Tetrachloroethene         | 2     | 2.00   | ug/L | 100 |     |      | 70 (70) | 130 (130)      |     |
|               | 1,1,1,2-Tetrachloroethane | 2     | 1.90   | ug/L | 95  |     |      | 70 (70) | 130 (130)      |     |