

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2377

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW524.2

Datafile : VU063433.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VU0624WBSD01	Vinyl Chloride	2	1.70	ug/L	85	0		70 (70)	130 (130)	20 (20)
	1,1-Dichloroethene	2	2.20	ug/L	110	26	*	70 (70)	130 (130)	20 (20)
	1,1-Dichloroethane	2	2.10	ug/L	105	5		70 (70)	130 (130)	20 (20)
	cis-1,2-Dichloroethene	2	2.00	ug/L	100	0		70 (70)	130 (130)	20 (20)
	Benzene	2	2.10	ug/L	105	5		70 (70)	130 (130)	20 (20)
	1,2-Dichloroethane	2	2.00	ug/L	100	0		70 (70)	130 (130)	20 (20)
	Trichloroethene	2	2.00	ug/L	100	5		70 (70)	130 (130)	20 (20)
	1,1,2-Trichloroethane	2	2.00	ug/L	100	5		70 (70)	130 (130)	20 (20)
	Tetrachloroethene	2	2.00	ug/L	100	0		70 (70)	130 (130)	20 (20)
	1,1,1,2-Tetrachloroethane	2	2.00	ug/L	100	5		70 (70)	130 (130)	20 (20)