

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2377 SAS No.: Q2377 SDG NO.: Q2377
Lab File ID: VU063430.D Date Analyzed: 06/24/2025
Instrument ID: MSVOA_U Time Analyzed: 10:17
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	56764	6.10	0	0.00	0	0.00
UPPER LIMIT	73793.2	6.603	0		0	
LOWER LIMIT	39734.8	5.603	0		0	
EPA SAMPLE NO.						
PW-B6-L66-061925	46278	6.11	0	0.00	0	0.00
TB01-061925	40746	6.11	0	0.00	0	0.00
VU0624WBL01	48463	6.11	0	0.00	0	0.00
VU0624WBS01	50423	6.10	0	0.00	0	0.00
VU0624WBSD01	48924	6.10	0	0.00	0	0.00

IS1 = Fluorobenzene

IS2 =

IS3 =

AREA UPPER LIMIT = +30% of internal standard area

AREA LOWER LIMIT = -30% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.