

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2377</u>
Lab File ID:	<u>VU063430.D</u>	Date Analyzed:	<u>06/24/2025</u>
Instrument ID:	<u>MSVOA_U</u>	Time Analyzed:	<u>10:17</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
PW-B6-L66-061925	0	0.00				
TB01-061925	0	0.00				
VU0624WBL01	0	0.00				
VU0624WBS01	0	0.00				
VU0624WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +30% of internal standard area
 AREA LOWER LIMIT = -30% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.