

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2378 SAS No.: Q2378 SDG NO.: Q2378

EPA Sample No.: SSTD020372 Date Analyzed: 06/24/2025

Lab File ID: BN037367.D Time Analyzed: 09:57

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	456082	16.963	504358	21.198	567860	24.045
UPPER LIMIT	912164	17.463	1008720	21.698	1135720	24.545
LOWER LIMIT	228041	16.463	252179	20.698	283930	23.545
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	348862	16.97	323454	21.20	352920	24.05

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.