

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2378 SAS No.: Q2378 SDG NO.: Q2378
 Lab File ID: VX046784.D Date Analyzed: 06/20/2025
 Instrument ID: MSVOA_X Time Analyzed: 15:49
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	86497	12.018				
UPPER LIMIT	172994	12.518				
LOWER LIMIT	43248.5	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	136509	12.02				
STORAGE-BLANK-WATER-REF-5	88541	12.02				
STORAGE-BLANK-WATER-REF-4	133303	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	86424	12.02				
VX0620WBL02	117754	12.02				
VX0620WBS02	99647	12.02				
VX0620WBSD02	94174	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.