

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q2390 SAS No.: Q2390 SDG No.: Q2390

Instrument ID: MSVOA\_L Calibration Date/Time: 06/23/2025 08:41

Lab File ID: VL042663.D Init. Calib. Date(s): 05/29/2025 05/29/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:24

GC Column: RTX-1 ID: 0.32 (mm)

| COMPOUND                | RRF   | RRF010 | MIN<br>RRF | %D    | MAX%D |
|-------------------------|-------|--------|------------|-------|-------|
| Vinyl Chloride          | 0.496 | 0.466  |            | -6.05 | 30    |
| 1,1-Dichloroethene      | 0.416 | 0.439  |            | 5.53  | 30    |
| cis-1,2-Dichloroethene  | 1.370 | 1.359  |            | -0.8  | 30    |
| 1,1,1-Trichloroethane   | 1.990 | 1.974  |            | -0.8  | 30    |
| Trichloroethene         | 0.419 | 0.429  |            | 2.39  | 30    |
| Tetrachloroethene       | 0.368 | 0.360  |            | -2.17 | 30    |
| 1-Bromo-4-Fluorobenzene | 0.750 | 0.795  |            | 6     | 30    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.