

Initial and Continuing Calibration Verification

Client: Earth Engineering Inc.

SDG No.: Q2392

Project: Thomas McGoven

RunNo.: LB136259

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	90-110	06/25/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	06/25/2025