

LB136362


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

44256



A= 0.0000e+000

B= 2.2579e-004

C= 4.4561e-002

Rho= 0.9995670

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.048	0.048	16	0.000	16					
0.2	0.200	0.223	0.023	791	0.0 %	791					
2.5	2.500	2.344	-0.156	10182	0.0 %	10182					
5.0	5.000	5.167	0.167	22685	0.0 %	22685					
7.5	7.500	7.381	-0.119	32491	0.0 %	32491					
10.0	10.000	10.038	0.038	44258	0.0 %	44258					

0/00
12
46
3
2
0

LB136362 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jun 2025 09:48:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7471B	PPB	26 Jun 2025 09:50:30	S	Std
0.2 - 1	50.2	791	-	0.2000	7471B	PPB	26 Jun 2025 09:52:47	S	Std
2.5 - 1	52.5	10182	-	2.5000	7471B	PPB	26 Jun 2025 09:55:04	S	Std
5.0 - 1	55	22685	-	5.0000	7471B	PPB	26 Jun 2025 09:57:21	S	Std
7.5 - 1	57.5	32491	-	7.5000	7471B	PPB	26 Jun 2025 09:59:39	S	Std
10.0 - 1	57.5	44258	-	10.0000	7471B	PPB	26 Jun 2025 10:01:58	S	Std
ICV06 - 1	ICV06	17222	3.9332	-7471B	PPB	26 Jun 2025 10:05:00	U	SMPL	
ICB06 - 1	ICB06	-266	-0.0155	-7471B	PPB	26 Jun 2025 10:07:15	U	SMPL	
CCV16 - 1	CCV16	22652	5.1592	-7471B	PPB	26 Jun 2025 10:09:34	U	SMPL	
CCB16 - 1	CCB16	-277	-0.0180	-7471B	PPB	26 Jun 2025 10:11:49	U	SMPL	
CRA - 1	CRA	774	0.2193	-7471B	PPB	26 Jun 2025 10:14:10	U	SMPL	
HighStd - 1	HighStd	44570	10.1081	-7471B	PPB	26 Jun 2025 10:16:25	U	SMPL	
ChkStd - 1	ChkStd	31462	7.1484	-7471B	PPB	26 Jun 2025 10:18:40	U	SMPL	
PB168616BL - 1	PBS	-388	-0.0430	-7471B	PPB	26 Jun 2025 10:21:10	U	SMPL	
PB168616BS - 1	LCSS	17080	3.9011	-7471B	PPB	26 Jun 2025 10:23:32	U	SMPL	
Q2392-01 - 1	S-1	759	0.2159	-7471B	PPB	26 Jun 2025 10:25:49	U	SMPL	
Q2393-01 - 1	PARK AVE -1	1503	0.3839	-7471B	PPB	26 Jun 2025 10:28:08	U	SMPL	
Q2393-02 - 1	PARK AVE -2	2789	0.6743	-7471B	PPB	26 Jun 2025 10:30:25	U	SMPL	
Q2393-05 - 1	PARK AVE -3	1590	0.4036	-7471B	PPB	26 Jun 2025 10:32:43	U	SMPL	
Q2393-07 - 1	PARK AVE -4	2240	0.5503	-7471B	PPB	26 Jun 2025 10:34:59	U	SMPL	
CCV17 - 1	CCV17	24003	5.4642	-7471B	PPB	26 Jun 2025 10:37:15	U	SMPL	
CCB17 - 1	CCB17	-244	-0.0105	-7471B	PPB	26 Jun 2025 10:39:31	U	SMPL	
Q2393-09 - 1	PARK AVE -5	1539	0.3921	-7471B	PPB	26 Jun 2025 10:41:55	U	SMPL	
Q2393-11 - 1	PARK AVE -6	1489	0.3808	-7471B	PPB	26 Jun 2025 10:44:10	U	SMPL	
Q2394-01 - 1	MH-K/L	520	0.1620	-7471B	PPB	26 Jun 2025 10:46:26	U	SMPL	
Q2398-03 - 1	M00-25-0179	306	0.1137	-7471B	PPB	26 Jun 2025 10:48:42	U	SMPL	
Q2398-06 - 1	ARS20-0036	-1	0.0443	-7471B	PPB	26 Jun 2025 10:50:57	U	SMPL	
Q2399-01 - 1	TP-13	2828	0.6831	-7471B	PPB	26 Jun 2025 10:53:14	U	SMPL	
Q2399-05 - 1	EP-7	1504	0.3842	-7471B	PPB	26 Jun 2025 10:55:30	U	SMPL	
Q2400-01 - 1	B-156-SB01	43668	9.9044	-7471B	PPB	26 Jun 2025 10:57:48	U	SMPL	
Q2400-02 - 1	B-134-SB01	629	0.1866	-7471B	PPB	26 Jun 2025 11:00:05	U	SMPL	
Q2403-03 - 1	LAW-25-0093	234	0.0974	-7471B	PPB	26 Jun 2025 11:02:50	U	SMPL	
CCV18 - 1	CCV18	22795	5.1915	-7471B	PPB	26 Jun 2025 11:07:10	U	SMPL	
CCB18 - 1	CCB18	-255	-0.0130	-7471B	PPB	26 Jun 2025 11:09:25	U	SMPL	
Q2403-07 - 1	ARS 20-006	2282	0.5598	-7471B	PPB	26 Jun 2025 11:11:49	U	SMPL	
Q2403-07DUP - 1	ARS 20-006DUP	1941	0.4828	-7471B	PPB	26 Jun 2025 11:14:04	U	SMPL	
Q2403-07MS - 1	ARS 20-006MS	19551	4.4590	-7471B	PPB	26 Jun 2025 11:16:20	U	SMPL	
Q2403-07MSD - 1	ARS 20-006MSD	18050	4.1201	-7471B	PPB	26 Jun 2025 11:18:36	U	SMPL	
Q2405-01 - 1	MH-M/H	9	0.0466	-7471B	PPB	26 Jun 2025 11:20:55	U	SMPL	
Q2403-07LX5 - 1		413	0.1378	-7471B	PPB	26 Jun 2025 11:23:14	U	SMPL	
Q2403-07A - 1		18810	4.2917	-7471B	PPB	26 Jun 2025 11:25:30	U	SMPL	
CCV19 - 1	CCV19	21880	4.9849	-7471B	PPB	26 Jun 2025 11:27:47	U	SMPL	
CCB19 - 1	CCB19	-263	-0.0148	-7471B	PPB	26 Jun 2025 11:30:07	U	SMPL	