

Continuing Calibration Report for SequenceID : FC062525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069277.D

Aliphatic C9-C12	8858286.000	60.000	3.306	6.605	147638.100	146679.627	-0.653
Aliphatic C12-C16	6280070.000	40.000	6.606	10.009	157001.750	159210.532	1.387
Aliphatic C16-C21	8934077.000	60.000	10.010	13.380	148901.283	154434.323	3.583
Aliphatic C21-C28	10815098.000	80.000	13.381	17.044	135188.725	133966.368	-0.912
Aliphatic C28-C40	12065799.000	120.000	17.045	22.027	100548.325	92124.417	-9.144
Aliphatic EPH	46953330.000	360.000	3.306	22.027	130425.917	128354.160	-1.614