

Continuing Calibration Report for SequenceID : FC062525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069286.D

Aliphatic C9-C12	8996087.000	60.000	3.306	6.605	149934.783	146679.627	-2.219
Aliphatic C12-C16	6473761.000	40.000	6.606	10.009	161844.025	159210.532	-1.654
Aliphatic C16-C21	9199380.000	60.000	10.010	13.380	153323.000	154434.323	0.720
Aliphatic C21-C28	11016234.000	80.000	13.381	17.044	137702.925	133966.368	-2.789
Aliphatic C28-C40	12362137.000	120.000	17.045	22.027	103017.808	92124.417	-11.825
Aliphatic EPH	48047599.000	360.000	3.306	22.027	133465.553	128354.160	-3.982